

MIT Open Access Articles

Calibration of high voltages at the ppm level by the difference of [^{83m}Kr conversion electron lines at the KATRIN experiment

The MIT Faculty has made this article openly available. **Please share** how this access benefits you. Your story matters.

Citation: Arenz, M., et al. "Calibration of High Voltages at the Ppm Level by the Difference of [^{83m}Kr Conversion Electron Lines at the KATRIN Experiment." The European Physical Journal C, vol. 78, no. 5, May 2018. © 2018 The Authors

As Published: <https://doi.org/10.1140/epjc/s10052-018-5832-y>

Publisher: Springer Berlin Heidelberg

Persistent URL: <http://hdl.handle.net/1721.1/116276>

Version: Final published version: final published article, as it appeared in a journal, conference proceedings, or other formally published context

Terms of use: Creative Commons Attribution



Calibration of high voltages at the ppm level by the difference of $^{83\text{m}}\text{Kr}$ conversion electron lines at the KATRIN experiment

M. Arenz¹, W.-J. Baek², M. Beck³, A. Beglarian⁴, J. Behrens², T. Bergmann⁴, A. Berlev⁵, U. Besserer⁶, K. Blaum⁷, T. Bode^{8,9}, B. Bornschein⁶, L. Bornschein¹⁰, T. Brunst^{8,9}, N. Buzinsky¹¹, S. Chilingaryan⁴, W. Q. Choi², M. Deffert², P. J. Doe¹², O. Dragoun¹³, G. Drexlin², S. Dyba¹⁴, F. Edzards^{8,9}, K. Eitel¹⁰, E. Ellinger¹⁵, R. Engel¹⁰, S. Enomoto¹², M. Erhard², D. Eversheim¹, M. Fedkevych¹⁴, S. Fischer⁶, J. A. Formaggio¹¹, F. M. Fränkle¹⁰, G. B. Franklin¹⁶, F. Friedel², A. Fulst¹⁴, W. Gil¹⁰, F. Glück¹⁰, A. Gonzalez Ureña¹⁷, S. Grohmann⁶, R. Grössle⁶, R. Gumbsheimer¹⁰, M. Hackenjos⁶, V. Hannen¹⁴, F. Harms², N. Haußmann¹⁵, F. Heizmann², K. Helbing¹⁵, W. Herz⁶, S. Hickford¹⁵, D. Hilk², D. Hillesheimer⁶, M. A. Howe^{18,19}, A. Huber², A. Jansen¹⁰, J. Kellerer², N. Kernert¹⁰, L. Kippenbrock¹², M. Kleesiek², M. Klein², A. Kopmann⁴, M. Korzeczek², A. Kovalík¹³, B. Krasch⁶, M. Kraus², L. Kuckert¹⁰, T. Lasserre^{9,20}, O. Lebeda¹³, J. Letnev²¹, A. Lokhov⁵, M. Machatschek², A. Marsteller⁶, E. L. Martin¹², S. Mertens^{8,9}, S. Mirz⁶, B. Monreal²², H. Neumann⁶, S. Niemes⁶, A. Off⁶, A. Osipowicz²¹, E. Otten³, D. S. Parno¹⁶, A. Pollithy^{8,9}, A. W. P. Poon²³, F. Priester⁶, P. C.-O. Ranitzsch¹⁴, O. Rest^{14,a}, R. G. H. Robertson¹², F. Roccati^{8,10}, C. Rodenbeck², M. Röllig⁶, C. Röttele², M. Ryšavý¹³, R. Sack¹⁴, A. Saenz²⁴, L. Schimpf², K. Schösser¹⁰, M. Schösser⁶, K. Schönung⁷, M. Schrank¹⁰, H. Seitz-Moskaliuk², J. Sentkerestiová¹³, V. Sibille¹¹, M. Slezák^{8,9}, M. Steidl¹⁰, N. Steinbrink¹⁴, M. Sturm⁶, M. Suchopar¹³, M. Suesser⁶, H. H. Telle¹⁷, L. A. Thorne¹⁶, T. Thümmler¹⁰, N. Titov⁵, I. Tkachev⁵, N. Trost¹⁰, K. Valerius¹⁰, D. Vénos¹³, R. Vianden¹, A. P. Vizcaya Hernández¹⁶, M. Weber⁴, C. Weinheimer¹⁴, C. Weiss²⁵, S. Welte⁶, J. Wendel⁶, J. F. Wilkerson^{18,19,b}, J. Wolf², S. Wüstling⁴, S. Zadoroghny⁵

- ¹ Helmholtz-Institut für Strahlen- und Kernphysik, Rheinische Friedrich-Wilhelms Universität Bonn, Nussallee 14-16, 53115 Bonn, Germany
- ² Institute of Experimental Particle Physics (ETP), Karlsruhe Institute of Technology (KIT), Wolfgang-Gaede-Str. 1, 76131 Karlsruhe, Germany
- ³ Institut für Physik, Johannes Gutenberg-Universität Mainz, 55099 Mainz, Germany
- ⁴ Institute for Data Processing and Electronics (IPE), Karlsruhe Institute of Technology (KIT), Hermann-von-Helmholtz-Platz 1, 76344 Eggenstein-Leopoldshafen, Germany
- ⁵ Institute for Nuclear Research of Russian Academy of Sciences, 60th October Anniversary Prospect 7a, 117312 Moscow, Russia
- ⁶ Institute for Technical Physics (ITeP), Karlsruhe Institute of Technology (KIT), Hermann-von-Helmholtz-Platz 1, 76344 Eggenstein-Leopoldshafen, Germany
- ⁷ Max-Planck-Institut für Kernphysik, Saupfercheckweg 1, 69117 Heidelberg, Germany
- ⁸ Max-Planck-Institut für Physik, Föhringer Ring 6, 80805 Munich, Germany
- ⁹ Technische Universität München, James-Franck-Str. 1, 85748 Garching, Germany
- ¹⁰ Institute for Nuclear Physics (IKP), Karlsruhe Institute of Technology (KIT), Hermann-von-Helmholtz-Platz 1, 76344 Eggenstein-Leopoldshafen, Germany
- ¹¹ Laboratory for Nuclear Science, Massachusetts Institute of Technology, 77 Massachusetts Ave, Cambridge, MA 02139, USA
- ¹² Center for Experimental Nuclear Physics and Astrophysics, and Dept. of Physics, University of Washington, Seattle, WA 98195, USA
- ¹³ Nuclear Physics Institute of the CAS, v. v. i, 250 68 Rež, Czech Republic
- ¹⁴ Institut für Kernphysik, Westfälische Wilhelms-Universität Münster, Wilhelm-Klemm-Str. 9, 48149 Münster, Germany
- ¹⁵ Department of Physics, Faculty of Mathematics and Natural Sciences, University of Wuppertal, Gauss-Str. 20, 42119 Wuppertal, Germany
- ¹⁶ Department of Physics, Carnegie Mellon University, Pittsburgh, PA 15213, USA
- ¹⁷ Universidad Complutense de Madrid, Instituto Pluridisciplinar, Paseo Juan XXIII no 1, 28040 Madrid, Spain
- ¹⁸ Department of Physics and Astronomy, University of North Carolina, Chapel Hill, NC 27599, USA
- ¹⁹ Triangle Universities Nuclear Laboratory, Durham, NC 27708, USA
- ²⁰ Commissariat à l'Énergie Atomique et aux Énergies Alternatives, Centre de Saclay, DRF/IRFU, 91191 Gif-sur-Yvette, France
- ²¹ University of Applied Sciences (HFD) Fulda, Leipziger Str. 123, 36037 Fulda, Germany
- ²² Department of Physics, Case Western Reserve University, Cleveland, OH 44106, USA
- ²³ Institute for Nuclear and Particle Astrophysics and Nuclear Science Division, Lawrence Berkeley National Laboratory, Berkeley, CA 94720, USA
- ²⁴ Institut für Physik, Humboldt-Universität zu Berlin, Newtonstr. 15, 12489 Berlin, Germany

²⁵ Project, Process, and Quality Management (PPQ), Karlsruhe Institute of Technology (KIT), Hermann-von-Helmholtz-Platz 1, 76344 Eggenstein-Leopoldshafen, Germany

Received: 20 February 2018 / Accepted: 22 April 2018
© The Author(s) 2018

Abstract The neutrino mass experiment KATRIN requires a stability of 3 ppm for the retarding potential at -18.6 kV of the main spectrometer. To monitor the stability, two custom-made ultra-precise high-voltage dividers were developed and built in cooperation with the German national metrology institute Physikalisch-Technische Bundesanstalt (PTB). Until now, regular absolute calibration of the voltage dividers required bringing the equipment to the specialised metrology laboratory. Here we present a new method based on measuring the energy difference of two $^{83\text{m}}\text{Kr}$ conversion electron lines with the KATRIN setup, which was demonstrated during KATRIN's commissioning measurements in July 2017. The measured scale factor $M = 1972.449(10)$ of the high-voltage divider K35 is in agreement with the last PTB calibration 4 years ago. This result demonstrates the utility of the calibration method, as well as the long-term stability of the voltage divider.

1 Introduction

Precision high voltages (HV) at the ppm level are required for many applications in science, e.g. for defining the kinetic energy of electrons in an electron cooler at storage rings [1] or for the precise determination of the energy of electrons in electrostatic retarding spectrometers or other analysers [2–4].

The Karlsruhe Tritium Neutrino (KATRIN) experiment [5] at the Karlsruhe Institute of Technology (KIT) (see Fig. 1) aims for a direct neutrino mass determination by a precise measurement of the tritium- β -decay spectrum near the endpoint. The expected sensitivity of the experiment is 0.2 eV/ c^2 at 90% C.L. [6]. Currently the Mainz- [7] and Troitsk- [8,9] neutrino mass experiments set upper limits on the neutrino mass of 2 eV/ c^2 .

In KATRIN, electrons are emitted from molecular tritium decaying in the windowless gaseous tritium source (WGTS) and are guided adiabatically by magnetic fields through the transport section. In this transport section, tritium is removed from the beamline by means of differential and cryogenic pumping. In the pre- and main spectrometers downstream from the transport section, the kinetic energy of the elec-

trons is analysed. In order to reach the desired sensitivity, the spectrometers need to provide a large acceptance angle for the emitted β -electrons and, in the case of the main spectrometer, a very good energy resolution as well. This is accomplished by operating the spectrometers as MAC-E filters [4], which are electrostatic retardation spectrometers (high energy filters) combined with magnetic adiabatic collimation to obtain large solid angle acceptance. Following the spectrometers, the focal-plane detector (FPD) [10] counts all electrons that have sufficient energy to pass both spectrometers. This FPD is a monolithic silicon p - i - n diode with 148 pixels.

One key requirement of the experiment is the stability of the retarding potential ($U_{\text{ret}} \approx -18.6$ kV) of the main spectrometer, which has to be maintained and monitored with a precision of 3 ppm (60 mV) over the measurement periods of 2 months. The knowledge of the absolute retarding potential would additionally allow a comparison of the tritium- β endpoint with the nuclear mass difference of ^3He and tritium ^3H determined with Penning traps [11–13].

Two independent approaches to monitor the high voltage are being pursued in order to ensure system redundancy. Firstly, since the HV cannot be measured directly with the required precision, a voltage divider is used to scale the retarding potential to $\lesssim 20$ V. The scaled retarding potential can then be determined with a commercial precision digital voltmeter (DVM). In this range, the voltage measurement can be calibrated against a 10 V reference, based on the Josephson effect, at the German national metrology institute Physikalisch-Technische Bundesanstalt (PTB). Since there is no commercial solution available with the required precision and stability, two custom-made ultra-precise high voltage dividers, named K35 [14] and K65 [15], were developed and built in cooperation with the PTB.

An HV divider is characterised by its scale factor M , which is defined as the ratio of input and output voltages. An absolute calibration with ppm-precision can usually be performed exclusively at metrology centres. The scale factor of the K35 was measured at PTB in 2013 to be

$$M_{\text{K35}}^{\text{PTB},2013} = 1972.4531(20). \quad (1)$$

Secondly, the HV is compared to a natural standard given by mono-energetic conversion electrons from the decay of $^{83\text{m}}\text{Kr}$. The parent isomer ^{83}Rb is implanted into a solid-state source [16] at the BONIS facility in Bonn [17] and used at the monitor spectrometer [18]. This third MAC-E

^a e-mail: oliver.rest@uni-muenster.de

^b Also affiliated with Oak Ridge National Laboratory, Oak Ridge, TN 37831, USA




Fig. 1 Experimental setup of the KATRIN experiment. The main components are (from left to right): calibration and monitoring rear section, windowless gaseous tritium source, transport section, pre-spectrometer, main spectrometer and focal plane detector

filter is connected to the HV of the main spectrometer. Due to solid-state and surface effects, such as electron energy losses in the source material and drifts of the work function, absolute calibration of the HV at the required precision is not possible with such a source. However, relative changes over a measurement period of up to several months can be monitored [19].

In July 2017, a calibration and measurement campaign with gaseous $^{83\text{m}}\text{Kr}$, injected into the WGTS from a $^{83\text{m}}\text{Kr}$ generating ^{83}Rb source [20], was performed with the complete KATRIN beamline [6]. With the well-known energies of mono-energetic conversion electron lines of this isotope, source properties of the WGTS and transmission properties of the spectrometers were investigated. Furthermore, the adiabatic transport of electrons from the source to the detector and the general alignment and functionalities of the complete system were tested [21, 22]. This measurement campaign also provided the opportunity to calibrate the K35 HV divider to the ppm-level by comparing two conversion electron lines. A similar HV calibration was previously performed using a condensed $^{83\text{m}}\text{Kr}$ source at the former Mainz neutrino mass experiment [23]. The main idea is to compare the kinetic energy of conversion electrons emitted from the same nuclear transition, but originating from different atomic shells. The systematic uncertainty of the nuclear transition energy cancels; the only remaining uncertainty, which is an order of magnitude lower, arises from the atomic binding energies. The measurements reported in [23] were limited by systematic corrections of the order 100 meV. These corrections, which are not precisely known, account for the final state effects of the decaying nucleus in a submonolayer of $^{83\text{m}}\text{Kr}$ on a highly oriented pyrolytic graphite (HOPG) substrate. Gaseous sources overcome the disadvantages of a condensed or solid-state source.

In this work, the calibration of the HV divider K35 with gaseous $^{83\text{m}}\text{Kr}$ is presented. The next section gives an overview of the calibration concept and the determination of the scale factor of the divider using $^{83\text{m}}\text{Kr}$ conversion electron line measurements. Subsequently, the results of the calibration measurements performed at KATRIN are reported.

2 Calibration of a HV divider with $^{83\text{m}}\text{Kr}$ conversion electrons

$^{83\text{m}}\text{Kr}$ decays via two cascaded transitions with gamma energies of 32151.6(5) and 9405.7(6) eV, respectively [24]. Both transitions decay dominantly by emission of conversion electrons instead of gamma radiation. In this work, only conversion electrons from the 32 keV transition are used.

The kinetic energy E_{kin} of a conversion electron from a $^{83\text{m}}\text{Kr}$ atom decaying freely in vacuum depends on the energy of the transition E_{γ} , the atomic binding energy E_{bin} , the nuclear recoil energies caused by the gamma¹ E_{rec}^{γ} , and the conversion electron $E_{\text{rec}}^{\text{ce}}$:

$$E_{\text{kin}} = E_{\gamma} - E_{\text{bin}} + E_{\text{rec}}^{\gamma} - E_{\text{rec}}^{\text{ce}}. \quad (2)$$

The binding energy depends on the atomic shell of the emitted electron. The values for E_{bin} used in this analysis were determined with X-ray and photoelectron spectroscopy measurements [2] to be 14327.26(4) eV for the K- and 1679.21(3) eV for the L_3 -subshell. The nuclear recoil energy of the conversion electrons $E_{\text{rec}}^{\text{ce}}$ can be calculated to be 0.120 eV for the K- and 0.207 eV for the L_3 -subshell for the 32 keV transition, both with negligible uncertainty.

$^{83\text{m}}\text{Kr}$ decays in the WGTS under ultra-high vacuum conditions. The conversion electrons are guided magnetically and adiabatically through the beamline to the main spectrometer, where an integrated spectrum is recorded by varying the retarding potential. With the HV divider K35 and a precision digital voltmeter (Fluke 8508A),² which measures the output voltage U_{DVM} of the K35, the corresponding retarding energy qU for a particle of charge q can be determined using

$$q \cdot U = q \cdot U_{\text{DVM}} \cdot M_{\text{K35}}, \quad (3)$$

¹ The nuclear recoil energy of the gamma transition E_{rec}^{γ} enters, since the nuclear transition energy ΔE_{fi} and the tabulated gamma energy E_{γ} differ by this nuclear recoil energy: $\Delta E_{\text{fi}} - E_{\text{rec}}^{\gamma} = E_{\gamma}$.

² The DVM was calibrated with a PTB-calibrated 10 V reference device (Fluke 732A).

where $q = -e$. The transmission condition for electrons to pass the main spectrometer is given by

$$E_{\text{kin}} \geq q \cdot U_{\text{DVM}} \cdot M_{\text{K35}} - \Delta\Phi - q \cdot U_{\text{pot.dec.}} \tag{4}$$

Since the retarding voltage is applied between the Fermi energies of the source tube and the spectrometer electrode system, $\Delta\Phi$ describes the correction for the difference between the work functions of the two materials. Due to the large diameter of the main spectrometer of 10 m and the wire electrode covering the inner surface [25], the retarding potential across the analysing plane is not perfectly equal to the applied potential and shows a radial dependence. Therefore, a pixel-wise correction for the potential $U_{\text{pot.dec.}}$ has been used. This correction amounts to about 2.25 V for the 40 innermost detector pixels with a r.m.s. value of less than 60 mV and scales nearly linearly in radial direction. The average difference of this correction over all these pixels amounts to 9 mV between the HV settings for the K-32 and the L₃-32 measurement.

The so-called transmission edge is a special case where the kinetic energy of the electrons equals the right-hand side of Eq. 4. Using Eqs. 2 and 4, the scale factor of the HV divider is then given as

$$M_{\text{K35}} = \frac{E_{\gamma} - E_{\text{bin}} + E_{\text{rec}}^{\gamma} - E_{\text{rec}}^{\text{ce}} + \Delta\Phi + q \cdot U_{\text{pot.dec.}}}{q \cdot U_{\text{DVM}}} \tag{5}$$

with U_{DVM} measured at the transmission edge. Following Eq. 5, the K35 could be calibrated by analysing just a single line position. However, the nuclear transition energy and the work function difference are not known to the desired ppm level. This limitation can be resolved using the energy difference of two conversion electron lines from the same gamma transition

$$M_{\text{K35}} = \frac{\Delta E_{\text{bin}} + \Delta E_{\text{rec}}^{\text{ce}} + q \cdot \Delta U_{\text{pot.dec.}}}{q \cdot \Delta U_{\text{DVM}}} \tag{6}$$

so that E_{γ} and $\Delta\Phi$ ³ are eliminated from the equation.⁴ Since K35 has a negligible voltage dependency of 0.03 ppm/kV [14], we assume a constant scale factor for the HV settings of the K-32 and the L₃-32 measurement. The differences of the binding and recoil energies

$$\Delta E_{\text{bin}} = E_{\text{bin}}^{\text{L}_3} - E_{\text{bin}}^{\text{K}} \tag{7}$$

³ The work functions of the source and spectrometer should be constant on the time scale of the measurements.

⁴ With a very different technique [26], collinear laser spectroscopy on Doppler-shifted ions, an absolute high voltage calibration has been reported recently. This novel method profits from eliminating systematics by performing two measurements at two different Doppler-shifts similarly to the method reported here.

$$\Delta E_{\text{rec}}^{\text{ce}} = E_{\text{rec}}^{\text{ce, L}_3\text{-32}} - E_{\text{rec}}^{\text{ce, K-32}} \tag{8}$$

add up to

$$\Delta E_{\text{bin}} + \Delta E_{\text{rec}}^{\text{ce}} = 12647.963(50)_{\text{sys}} \text{ eV}. \tag{9}$$

The potential correction $U_{\text{pot.dec.}}$ has been determined for every FPD pixel by an electric field calculation with the simulation software Kassiopeia [27].

In order to determine the individual line energy positions, the observed integral spectrum was fitted with MINUIT [28]. The fit function consists of a Lorentzian with free amplitude a , width Γ and the energy $E(\text{K-32})$ or $E(\text{L}_3\text{-32})$, convolved with the transmission function $T(E, U_{\text{DVM}})$ of the main spectrometer:

$$T(E, U_{\text{DVM}}) = \begin{cases} 0 & \text{for } E - qU \leq 0 \\ \frac{1 - \sqrt{1 - \frac{E - qU}{E} \frac{B_s}{B_a} \frac{2}{\gamma + 1}}}{1 - \sqrt{1 - \frac{B_s}{B_m}}} & \text{for } 0 < E - qU < \Delta E_{\text{trans}} \\ 1 & \text{for } E - qU \geq \Delta E_{\text{trans}} \end{cases} \tag{10}$$

and a constant background term b [6]. Here we used the abbreviation $U = U_{\text{DVM}} \cdot M_{\text{K35}}$ from Eq. 3. Relativistic corrections are included in Eq. 10 using the Lorentz factor γ of the electron. The width of the transmission function $\Delta E_{\text{trans}} = E \cdot \frac{B_a}{B_m} \cdot \frac{\gamma + 1}{2}$ is calculated from the energy of the electrons and the ratio of the magnetic flux densities in the analysing plane ($B_a = 0.268$ mT) and that at the exit of the spectrometer ($B_m = 4.20$ T).

Equation 10 is derived assuming an isotropically emitting source. In the case that the source magnetic field ($B_s = 2.52$ T) is lower than the maximum magnetic field encountered by the electrons on their way to the detector, the maximum amplitude of the transmission function is limited by magnetic reflection of the electrons with emission angles that exceed a cut-off angle given by

$$\theta_{\text{start}}^{\text{max}} = \arcsin\left(\sqrt{\frac{B_s}{B_m}}\right) \approx 50.8^\circ. \tag{11}$$

The final fit function for the K-32 line is

$$f(E) = \int_{qU_{\text{DVM}}M_{\text{K35}}}^{\infty} \frac{a/\pi \cdot \Gamma/2}{(E(\text{K-32}) - E')^2 + \Gamma^2/4} \cdot T'(E', U_{\text{DVM}}) dE' + b, \tag{12}$$

where the modified transmission function $T'(E', U_{\text{DVM}})$ contains three additional corrections: Firstly, the tempera-




Fig. 2 Averaged K-32 (left) and L₃-32 (right) data of the innermost 40 detector pixels with a four-parameter line fit to visualize the more complex constrained fit of the 40 individual pixels (see text). The denoted line position does not include the work function difference $\Delta\Phi$ between

the source and the spectrometer, which is not known to ppm precision, or other systematic uncertainties. The upper abscissa provides the corresponding energies according to Eqs. 1 and 3. The panels below the fits display the normalised residuals

ture of the ^{83m}Kr gas in the WGTS of 100 K leads to a thermal Gaussian broadening of the conversion line. Secondly, a high voltage ripple of the retarding potential was observed throughout the measurements [21]. The ripple had a nearly sinusoidal shape with a frequency of 50 Hz and an amplitude of 187 mV for the K-32 line and of 208 mV for the L₃-32 line. The Gaussian broadening and the recorded ripple signal are convolved with the transmission function in the fit. Thirdly, the shape of the transmission function will be modified by synchrotron radiation losses,⁵ increasing its width ΔE_{trans} by about 3% (2%) for the K-32 (L₃-32) line.

3 Calibration results for the HV divider K35

During the KATRIN calibration and measurement phase in July 2017, the energy of all conversion electron lines of the gaseous ^{83m}Kr source were measured. The K-32 and L₃-32

lines were used to calibrate the high-voltage divider K35 as described in Sect. 2.

In this work, a combined analysis of the 40 innermost detector pixels (out of 148 pixels in total) was performed to obtain high statistics while avoiding increased systematic uncertainties at larger beam radii. Each detector pixel was treated with its corresponding potential correction $U_{pot.dec.}$. For illustration, the average of all K-32 and L₃-32 conversion electron data of these 40 innermost detector pixels has been calculated and fitted, as shown in Fig. 2.

The good agreement between data and the fit model can be seen in the residuals as well as in the reduced χ^2 values of the fits.

For the final result, we avoid averaging the pixel-dependent $U_{pot.dec.}$ values by performing a combined 82-parameter fit⁶ of the data from the 40 innermost detector pixels, leading to the results shown in Table 1.

The results from Table 1 yield a voltage difference of

$$\Delta U_{DVM} = 6.412315(15)_{stat}(15)_{sys} \text{ V.} \tag{13}$$

⁵ The synchrotron radiation affects the transversal energy E_{\perp} (i.e. in the motion direction transversal of the magnetic field B) with the power loss $\dot{E}_{\perp} = -\frac{e^4 \beta^2 \gamma^2}{6\pi \epsilon_0 m_e^2 c} \cdot B^2$, where γ is the relativistic factor and $\beta = v_{\perp}/c$ is the velocity. In the non-relativistic case, the power loss amounts to $\dot{E}_{\perp} = -\frac{0.39}{[T^2 s]} \cdot E_{\perp} \cdot B^2$. Hence electrons emitted with high angles will be transmitted at lower retarding potentials, which results in a broadening of the transmission function.

⁶ In this combined fit we used common fit parameters for line width Γ and position $E(K-32)$ (or $E(L_3-32)$, respectively) but with separate fit parameters for amplitude a_i and background b_i for each pixel ($i = 1, \dots, 40$).

Table 1 Fit result for the combined analysis of the 40 innermost detector pixels. For the denoted line, the work function difference $\Delta\Phi$ between the source and the spectrometer is not known with sufficient precision and is not included. The same holds for other systematic uncertainties. We estimate the uncertainty of $\Delta\Phi$ to be of the order of a few 100 meV. As $\Delta\Phi$ drops out in the calculation of ΔU_{DVM} , this does not pose a problem for further analyses

parameter	K-32	L ₃ -32
U_{DVM} line position (V)	9.036768(12)	15.449083(9)
U_{DVM} line width (V)	0.00135(4)	0.00056(2)
χ^2/N_{dof}	$\frac{1131.91}{1158}=0.98$	$\frac{1257.51}{1198}=1.05$

In the evaluation of the systematic uncertainties associated with this measurement, we considered a $\pm 20\%$ variation of the high-voltage ripple amplitude and a $\pm 50\%$ uncertainty of the synchrotron-radiation correction. The systematic uncertainty of the synchrotron radiation was estimated very conservatively because we did not apply a pixel-wise correction. The assumed ± 5 meV uncertainty on the variation of $U_{\text{pot.dec.}}$ for the different conversion electron lines results in an uncertainty of $\pm 2.5 \mu\text{V}$ for ΔU_{DVM} (Eq. 3).

In the voltage determination with the DVM, we applied a 0.5 ppm uncertainty on the read value and a 0.2 ppm uncertainty on the full range of the device. These effects yield uncertainties of $8.5 \mu\text{V}$ ($11.7 \mu\text{V}$) for the K-32 (L₃-32) voltage reading, and $14.5 \mu\text{V}$ for ΔU_{DVM} .

Since the term $q \cdot \Delta U_{\text{pot.dec.}}$ was already absorbed in the fitted data, the scale factor can be determined simply by dividing Eq. 9 by Eq. 13:

$$M_{K35} = 1972.4488(45)_{\text{stat}}(91)_{\text{sys}} \approx 1972.449(10). \quad (14)$$

This result is in good agreement with the last calibration at PTB (Eq. 1) within the uncertainties. With a four-year interval between the two calibrations, the relative deviation amounts to $\Delta M/M = -2(5)$ ppm. This means that the stability of the scale factor is on the ppm-level per year or better, assuming a constant drift. For a typical KATRIN measurement period, which is partitioned in 2-month intervals, sub-ppm-stability can be assumed.

The uncertainty of 5 ppm of this new calibration method is dominated by the uncertainty of the difference of the atomic binding energies (relative uncertainty of 4 ppm). This could improve in the next years with more precise spectroscopic measurements or theoretical calculations. The combined relative statistical uncertainty of about 2 ppm can be improved by future measurements with higher statistics during calibration phases at KATRIN. The similarly large uncertainty of the voltage reading could be improved by measuring the two conversion lines in quick succession (~ 20 min.) to mitigate the temporal drift effect of the device.

This measurement has also demonstrated that the relative stability of the HV divider is better than 3 ppm in a 2-month interval, which significantly surpasses the design specifications.

4 Conclusion

In order to achieve the design sensitivity of $0.2 \text{ eV}/c^2$ in the neutrino mass measurement, the retarding potential of the main spectrometer of the KATRIN experiment has to be monitored with a precision of 3 ppm over measurement intervals of 2 months. The retarding voltage is measured with two custom-made ultra-precise HV dividers that have to be calibrated regularly. In the past, such calibrations could only be performed at the special metrology laboratories. In this work, a new calibration method is presented, which based on the energy difference of two conversion electron lines produced by the decay of $^{83\text{m}}\text{Kr}$. This method was previously applied with a condensed $^{83\text{m}}\text{Kr}$ source at the Mainz neutrino mass experiment, but surface and solid-state effects limited the attainable precision. Measurements with gaseous $^{83\text{m}}\text{Kr}$ at the KATRIN experiment are not affected by these effects, and allow the HV dividers to be calibrated with an uncertainty of < 5 ppm. We have shown in this paper that such precision is achievable. The measured scale factor of the divider K35 $M_{K35} = 1972.449(10)$ is in agreement with earlier PTB calibrations. The results demonstrate the stability and reliability of the K35 HV divider to sub-ppm-levels over the 2-month measurement intervals in KATRIN. This principle of determining the difference of two conversion electron lines with an electrostatic retardation spectrometer, e.g. of MAC-E-filter type, can be applied to other energy lines in other applications.

Acknowledgements We acknowledge the support of Helmholtz Association (HGF), Ministry for Education and Research BMBF (05A14VK2 and 05A17PM3), Helmholtz Alliance for Astroparticle Physics (HAP), and Helmholtz Young Investigator Group (VH-NG-1055) in Germany; Ministry of Education, Youth and Sport (CANAM-LM2011019, LTT18021), in cooperation with JINR Dubna (3+3 grants) in the Czech Republic; and the Department of Energy through Grants DEFG02-97ER41020, DE-FG02-94ER40818, DE-SC0004036, DEFG02-97ER41033, DE-FG02-97ER41041, DE-AC02-05CH11231, and DE-SC0011091 in the United States.

Open Access This article is distributed under the terms of the Creative Commons Attribution 4.0 International License (<http://creativecommons.org/licenses/by/4.0/>), which permits unrestricted use, distribution, and reproduction in any medium, provided you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made. Funded by SCOAP³.

References

1. J. Ullmann, Z. Andelkovic, C. Brandau, A. Dax, W. Geithner, C. Geppert, C. Gorges, M. Hammen, V. Hannen, S. Kaufmann, K. König, Y.A. Litvinov, M. Lochmann, B. Maaß, J. Meisner, T. Murböck, R. Sánchez, M. Schmidt, S. Schmidt, M. Steck, T. Stöhlker, R.C. Thompson, C. Trageser, J. Vollbrecht, C. Weinheimer, W. Nörtershäuser, Nat. Commun. **8**, 15484 EP (2017). <https://doi.org/10.1038/ncomms15484>
2. O. Dragoun, A. Špalek, F. Wuilleumier, Czechoslovak J. Phys. **54**(8), 833 (2004). <https://doi.org/10.1023/B:CJOP.0000038591.13369.e1>
3. P. van der Heide, *XPS instrumentation* (Wiley, Hoboken, 2011), pp. 27–60. <https://doi.org/10.1002/9781118162897.ch3>
4. A. Picard, H. Backe, H. Barth, J. Bonn, B. Degen, T. Edling, R. Haid, A. Hermanni, P. Leiderer, T. Loeken, A. Molz, R. Moore, A. Osipowicz, E. Otten, M. Przyrembel, M. Schrader, M. Steininger, C. Weinheimer, Nucl. Instr. Methods Phys. Res. Sect. B Beam Interact. Mater. Atoms **63**(3), 345 (1992). [https://doi.org/10.1016/0168-583X\(92\)95119-C](https://doi.org/10.1016/0168-583X(92)95119-C)
5. KATRIN collaboration, *KATRIN design report*. <https://www.katrin.kit.edu/publikationen/DesignReport2004-12Jan2005.pdf>
6. G. Drexlin, V. Hannen, S. Mertens, C. Weinheimer, Adv. High Energy Phys. **2013**, 293986 (2013). <https://doi.org/10.1155/2013/293986>
7. C. Kraus et al., Eur. Phys. J. C **40**, 447 (2005). <https://doi.org/10.1140/epjc/s2005-02139-7>
8. V.M. Lobashev, Nucl. Phys. A **719**, 153 (2003). [https://doi.org/10.1016/S0375-9474\(03\)00985-0](https://doi.org/10.1016/S0375-9474(03)00985-0)
9. V.N. Aseev et al., Phys. Rev. D **84**, 112003 (2011). <https://doi.org/10.1103/PhysRevD.84.112003>
10. J. Amsbaugh, J. Barrett, A. Beglarian, T. Bergmann, H. Bichsel, L. Bodine, J. Bonn, N. Boyd, T. Burritt, Z. Chaoui, S. Chilingaryan, T. Corona, P. Doe, J. Dunmore, S. Enomoto, J. Formaggio, F. Fraenkle, D. Furse, H. Gemmeke, F. Glueck, F. Harms, G. Harper, J. Hartmann, M. Howe, A. Kaboth, J. Kelsey, M. Knauer, A. Kopmann, M. Leber, E. Martin, K. Middleman, A. Myers, N. Oblath, D. Parno, D. Peterson, L. Petzold, D. Phillips, P. Renschler, R. Robertson, J. Schwarz, M. Steidl, D. Tcherniakhovski, T. Thmmmler, T.V. Wechel, B. VanDevender, S. Vcking, B. Wall, K. Wierman, J. Wilkerson, S. Wuestling, Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment **778**, 40 (2015). <https://doi.org/10.1016/j.nima.2014.12.116>
11. S. Streubel, T. Eronen, H.M.J. Ketter, M. Schuh, R.S. Van Dyck Jr., K. Blaum, Appl. Phys. **B114**, 137 (2014). <https://doi.org/10.1007/s00340-013-5669-x>
12. R.S. Van Dyck, D.L. Farnham, P.B. Schwinberg, Phys. Rev. Lett. **70**, 2888 (1993). <https://doi.org/10.1103/PhysRevLett.70.2888>
13. E.G. Myers, A. Wagner, H. Kracke, B.A. Wesson, Phys. Rev. Lett. **114**(1), 013003 (2015). <https://doi.org/10.1103/PhysRevLett.114.013003>
14. T. Thümmmler, R. Marx, C. Weinheimer, N. J. Phys. **11**, 103007 (2009). <https://doi.org/10.1088/1367-2630/11/10/103007>
15. S. Bauer, R. Berendes, F. Hochschulz, H.W. Ortjohann, S. Rosendahl, T. Thümmmler, M. Schmidt, C. Weinheimer, JINST **8**, P10026 (2013). <https://doi.org/10.1088/1748-0221/8/10/P10026>
16. M. Zboril, S. Bauer, M. Beck, J. Bonn, O. Dragoun, J. Jakubek, K. Johnston, A. Kovalik, E.W. Otten, K. Schloesser, M. Slezák, A. Spalek, T. Thuemmler, D. Venos, J. Zemlicka, C. Weinheimer, J. Instr. **8**(03), P03009 (2013). <http://stacks.iop.org/1748-0221/8/i=03/a=P03009>
17. M. Arenz, doctoral thesis, University of Bonn (2017). <http://hss.ulb.uni-bonn.de/2017/4930/4930.htm>
18. M. Erhard et al., JINST **9**, P06022 (2014). <https://doi.org/10.1088/1748-0221/9/06/P06022>
19. M. Slezák, PhD thesis, Nuclear Physics Institute Czech Academy of Sciences (2015). http://www.katrin.kit.edu/publikationen/phd-Martin_Slezak.pdf
20. J. Sentkerestiová, D. Vénos, M. Slezák, *Journal of Physics: Conference Series* **888**(1), 012072 (2017). <http://stacks.iop.org/1742-6596/888/i=1/a=012072>
21. M. Arenz, et al., (KATRIN collaboration), to be published (2018). <https://arxiv.org/abs/1802.04167>
22. M. Arenz, et al., (KATRIN collaboration), to be published (2018)
23. T. Thümmmler, doctoral thesis, University of Münster (2007). <https://miami.uni-muenster.de/Record/0e96c5e3-6f06-489e-a73f-5859f9be9cae7>
24. E. McCutchan, Nuclear Data Sheets **125**(Supplement C), 201 (2015). <https://doi.org/10.1016/j.nds.2015.02.002>
25. K. Valerius, Prog. Part. Nucl. Phys. **64**, 291 (2010). <https://doi.org/10.1016/j.pnpnp.2009.12.032>
26. J. Krämer, K. König, C. Geppert, P. Imgram, B. Maa, J. Meisner, E.W. Otten, S. Passon, T. Ratajczyk, J. Ullmann, W. Noerter-shaeuser, Metrologia **55**(2), 268 (2018). <http://iopscience.iop.org/article/10.1088/1681-7575/aaabe0>
27. D. Furse, S. Groh, N. Trost, M. Babutzka, J.P. Barrett, J. Behrens, N. Buzinsky, T. Corona, S. Enomoto, M. Erhard, J.A. Formaggio, F. Glueck, F. Harms, F. Heizmann, D. Hilke, W. Kaefer, M. Kleesiek, B. Leiber, S. Mertens, N.S. Oblath, P. Renschler, J. Schwarz, P.L. Slocum, N. Wandkowsky, K. Wierman, M. Zacher, New Journal of Physics **19**(5), 053012 (2017). <http://stacks.iop.org/1367-2630/19/i=5/a=053012>
28. F. James, *MINUIT—function minimization and error analysis* (CERN, Geneva, 1994)