

MIT Open Access Articles

Closed-loop feedback control for microfluidic systems through automated capacitive fluid height sensing

The MIT Faculty has made this article openly available. **Please share** how this access benefits you. Your story matters.

Citation: Soenksen, L. R. et al. "Closed-Loop Feedback Control for Microfluidic Systems through Automated Capacitive Fluid Height Sensing." *Lab on a Chip* 18, 6 (2018): 902–914 © 2018 Royal Society of Chemistry

As Published: <http://dx.doi.org/10.1039/C7LC01223C>

Publisher: Royal Society of Chemistry (RSC)

Persistent URL: <http://hdl.handle.net/1721.1/116892>

Version: Final published version: final published article, as it appeared in a journal, conference proceedings, or other formally published context

Terms of use: Creative Commons Attribution-NonCommercial 3.0 Unported




 Cite this: *Lab Chip*, 2018, 18, 902

Closed-loop feedback control for microfluidic systems through automated capacitive fluid height sensing†

 L. R. Soenksen, *^{ab} T. Kassis, ^{bc} M. Noh, ^{ab} L. G. Griffith^{ac} and D. L. Trumper*^{ab}

Precise fluid height sensing in open-channel microfluidics has long been a desirable feature for a wide range of applications. However, performing accurate measurements of the fluid level in small-scale reservoirs (<1 mL) has proven to be an elusive goal, especially if direct fluid-sensor contact needs to be avoided. In particular, gravity-driven systems used in several microfluidic applications to establish pressure gradients and impose flow remain open-loop and largely unmonitored due to these sensing limitations. Here we present an optimized self-shielded coplanar capacitive sensor design and automated control system to provide submillimeter fluid-height resolution (~250 μm) and control of small-scale open reservoirs without the need for direct fluid contact. Results from testing and validation of our optimized sensor and system also suggest that accurate fluid height information can be used to robustly characterize, calibrate and dynamically control a range of microfluidic systems with complex pumping mechanisms, even in cell culture conditions. Capacitive sensing technology provides a scalable and cost-effective way to enable continuous monitoring and closed-loop feedback control of fluid volumes in small-scale gravity-dominated wells in a variety of microfluidic applications.

 Received 16th November 2017,
 Accepted 2nd February 2018

DOI: 10.1039/c7lc01223c

rsc.li/loc

I. Introduction

Devices for precise fluid handling at the micro- and meso-scale, with characteristic lengths (λ) < 10 cm, are becoming increasingly common across a variety of fields, including biological, pharmaceutical and medical research.¹ Most microfluidic systems require at least one pump to produce flow, as well as open- or closed-channels or reservoirs to supply, store and collect fluid. Such devices can be categorized in terms of their primary pumping action as mechanical, chemical, surface-, force-field and mixed systems.^{1–3} Examples of mechanical pumping at these scales include syringe-pumps, microelectromechanical (MEM) pumps,⁴ magnetically actuated pumps,^{5,6} vacuum or pressure-driven pumps, peristaltic pumps and centrifugal pumps.^{1,7,8} Chemically-induced flow can be seen in osmotic and effervescence pumps, as well as systems using electrochemical reactions.^{1,9–11} Surface-effect pumps include all devices making use of spontaneous capillary forces and substrate wicking to passively induce fluid

flow.¹² Further, force-field pumps are usually defined as non-mechanical including electro-osmotic,^{3,13} electrowetting,¹⁴ poroelastic,¹⁵ electrodynamic, thermocapillary and gravity-driven pumps.^{1,3,16,17}

Common metrics for performance comparison in microfluidic applications include: operational fluid volumes, flow rates, directionality, persistence time, flow pattern control, fluid capacity, recirculation, optical accessibility, power requirements, heat transfer, and cost of implementation.¹ Furthermore, criteria such as imposing minimal mechanical, chemical and electrical stress during cell or particle handling, as well as providing material/surface stability against fouling or analyte absorption may also be important to consider.¹ Understanding the expected performance and inherent limitations associated with each fluid handling and pumping scheme is essential. For example, mechanical pumps connected to open wells or channels do not usually deliver robust steady-state flows for long periods of time as they are bounded by a specific input volume (*e.g.* syringe pumps), rely on pulsatile flow (*e.g.* peristaltic, pneumatic and MEM-based pumps),^{11,12,18–21} are transient in nature (*e.g.* gravity-driven,²² capillary-driven¹² and centrifugal pumps²³), or are prone to performance variation (*e.g.* fabrication errors, use-induced stress). Thus, it remains desirable to identify novel approaches to provide robust, scalable and cost-effective monitoring and control of both pumping performance and fluid volume allocation in microfluidic devices.

^a Department of Mechanical Engineering, MIT, Cambridge, MA, USA.
 E-mail: soenksen@mit.edu

^b Research Laboratory of Electronics, MIT, Cambridge, MA, USA

^c Department of Biological Engineering, MIT, Cambridge, MA, USA

† Electronic supplementary information (ESI) available. See DOI: 10.1039/c7lc01223c



Among the most attractive targets for improved microfluidic pumping control and volume tracking, we find a variety of open-channel systems such as multi-well cell-culture plates and organ-on-chip platforms.^{24,25} In these systems, open wells are usually intended to maintain constant fluid volumes while driven by passive or active recirculation and media oxygenation circuits that extract and replenish fluid from the system.²⁵ Variations in input and output pumping performance, as well as unpredictable variations in evaporation rates often lead to significant errors in desired fluid heights over relevant experimental timescales. In the particular case of cell culture applications where high flow-rates are required, fluid height errors in the order of millimeters may render otherwise healthy cultures nonviable after just a few hours of operation.²⁶ Moreover, if testing for drug toxicity and other effects related to compound concentration in such experiments, tight fluid volume tracking becomes an absolute necessity for analysis, which is something largely absent in open-channel fluidic devices used for this purpose.²⁷ Indeed, accurate fluid level monitoring capabilities at these scales has been a challenging goal, with current solutions being mostly incompatible with sterile or contact-sensitive applications, or requiring bulky and expensive instrumentation to achieve.²⁸ This situation has not only limited the reproducibility of experiments performed in a variety of microfluidic platforms, but has also hindered the use of feedback control to improve the robustness of these systems.

Traditional fluid-level sensing technologies such as mechanical floaters, displacers, bubble, magnetic level gauges, magnetostrictive level transmitters and ultrasonic distance sensors are all currently inadequate or unreliable at the scale of most micro- and mesofluidic devices.²⁹ More recent options such as in-line flow sensors and differential pressure transducers have become increasingly smaller and sensitive over the past years, and are now used in several microfluidic applications.^{30–32} Nonetheless, these and other available sensing techniques are limited in that they require some degree of fluid contact, which may be undesirable in many sterile biological applications. Non-contact fluid-level sensing technologies; on the other hand, have only become popular in a limited number of large-scale industrial applications including monitoring and handling of large volumes of hazardous fluids. From direct visualization of fluid height (*e.g.* using optical monitoring through translucent windows), to time-of-flight (TOF) measurements (*e.g.* ultrasonic, laser and radar transmitters), and capacitive fluid sensing;³³ non-contact fluid-level sensors are now an increasingly common strategy to control macrofluidic systems where relevant height changes happen in the order of tens of centimeters. From these non-contact options, capacitive fluid sensing is perhaps the most attractive due to scalability and low cost. In the case of microfluidic applications, several groups have previously demonstrated the use of capacitive sensing for binary detection of droplets³⁴ and droplet volume quantification.^{17,35} Furthermore, other research groups have used this technology to develop contact-based in-line pressure sen-

sors,³⁵ assess composition of liquids³⁶ and even small-scale flow directionality sensors.³⁷ While these developments show the general versatility and ease of integration of capacitive sensing in fluidic systems; to our knowledge, the use of capacitive fluid level-sensing for the control of open-well chambers in microfluidics has yet to be demonstrated. Here we present an optimized non-contact capacitive fluid level-sensing technology with submillimeter resolution to provide closed-loop control of a gravity-driven pump for cell culture as a proof-of-concept to show its use in a range of microfluidic systems.

II. Materials and methods

II.a Feedback-controlled gravity-driven pump setup

A schematic of the assembled testing setup developed to demonstrate the capabilities of our capacitive fluid-level sensor can be seen in Fig. 1A. The elements depicted in this system include a capacitive-fluid sensor, a hydrostatic chamber, a microcontroller unit, a bidirectional pump, a microfluidic device and a recirculation container, as well as tubing and connectors. In this testing setup, the fluid level in the primary fluid reservoir is set to the desired position (H1) to act as a hydrostatic chamber, which drives gravity-driven flow through a meso- and/or microfluidic chip. This chip is located inside a device holder to be connected to the fluidic




Fig. 1 A) Schematic of the proposed closed-loop controlled gravity-driven microfluidic setup with capacitive fluid-level sensing. In this system, the target microfluidic chip is driven by only one hydrostatic pressure chamber, although multiple chambers could be used. B) Simplified closed-loop feedback-control block diagram of the gravity-driven setup. Reference input $x(t)$ is user-defined and compared to the measured state (*i.e.* fluid height, pressure) to calculate the error $e(t)$. Fluid I/O's refer to the active bidirectional supply pumps that control the height in the primary chamber. $C(s)$, $P(s)$ and $G(s)$ are the controller, fluidic plant and sensor transfer functions respectively.



circuit *via* standard tubing and microfluidic connectors. The height difference (ΔH) with respect to the chamber outlet produces a hydrostatic pressure (P_{in}) at the inlet of the connected microfluidic channel such that $P_{in} = \rho g \Delta H$; where $\rho \approx 1 \text{ g cm}^{-3}$ (fluid density for water) and $g \approx 9.81 \text{ m s}^{-2}$. The flow rate imposed through the microfluidic device (Q) is then related to the pressure gradient between the inlet and outlet of the microfluidic channel (ΔP) as well as its fluidic resistance (R) such that $Q = \Delta P/R$. During operation, the fluid level in this primary reservoir is continuously monitored using the non-contact fluid-level sensor. The recorded fluid level is then fed to a microcontroller unit (MCU) implementing a closed-loop algorithm (Fig. 1B).

This system takes a user-defined reference input signal $x(t)$ (*i.e.* height, pressure, flow rate) and compares it to the measured state to produce an error signal $e(t)$. The generated error feeds a control law that actuates a bidirectional pump capable of actively increasing or decreasing the fluid height in the primary fluid reservoir to track the reference input $x(t)$. Standard tubing, connectors and a secondary container complete the fluidic circuit, all of which allow for stand-alone operation. A prototype of this system, including required electronic components and other structural elements (*e.g.* controller box, sensor/chamber holder, microfluidic device holder) can be seen in Fig. 2A. More detail on the design and fabrication of each of the components and modules is presented in further sections.

II.a.1 Hydrostatic fluid chamber. Fig. 2C shows the assembled hydrostatic chamber with capacitive sensing used in the feedback-controlled gravity-driven setup. In this assembly, a

sterile 1 mL Plastipak™ graduated syringe barrel (Becton Dickinson, Rutherford, USA) is oriented vertically and connected to a 1/16" polypropylene barbed quick-turn coupling socket (McMaster Carr, Robbinsville, USA) within a 3D printed structure to form a biocompatible and sterile reservoir with dimensions ($H_{total} = 60 \text{ mm}$, ID = 4.78 mm). A 0.22 μm pore Fisherbrand™ filter (Thermo Fisher Scientific, Waltham, USA) is placed on top of the assembly to allow for air flow while maintaining sterility. This hydrostatic chamber is connected to the pumping and recirculation circuits through 1.59 mm (1/16") ID polypropylene tubing and a compatible nylon tube-to-tube wye connector (McMaster Carr, Robbinsville, USA).

Gravity-driven flow is determined by the height of the fluid column and the downstream resistance of the system. The microfluidic device is connected using standard tubing and located within the device holder (Fig. 2A) to reliably control its vertical position with respect to the bottom of the fluid column. The non-contact capacitive fluid-level sensor is in close proximity to the fluid ($\sim 2 \text{ mm}$) in the 3D printed structure (Fig. 2B) so as to monitor fluid height. Additional details on the components used for the assembly of this monitored hydrostatic chamber can be found in Fig. S1 of the ESI.†

II.a.2 Capacitive fluid-level sensor for microfluidics. Our optimized capacitive fluid-level sensor for microfluidic applications consists of self-shielded coplanar electrodes connected to an AD7746 24 bit $\Sigma\text{-}\Delta$ capacitance-to-digital converter in differential mode (Analog Devices, Norwood, USA). Fig. 3A illustrates the layout of this sensor in conjunction with the monitored fluid reservoir. In our design, two pairs of excitation




Fig. 2 A) Fabricated testing setup used to perform the validation experiments of height-based closed-loop control of open-well microfluidic systems. B) Detail of embedded non-contact capacitive fluid-level sensor with custom-designed electrodes. C) Assembled gravity driven reservoir with embedded capacitive sensing and a single upper air filter (0.22 μm pore size) for sterile use in cell culture applications.






Fig. 3 Design of the capacitive fluid-level sensor for microfluidic applications. A) Planes filled with diagonal patterns are excitation electrodes, while planes with dotted patterns are sensing and reference electrodes. The yellow plane is a middle polyimide dielectric layer, while the blue cylinder represents the monitored fluid reservoir in experimental setup. B) Shows the sensing, reference and excitation electrodes connected to their respective terminals on the CDC chip. C) Cross-sectional detail of charge distribution and fringing capacitance fields (red arrows) among sensing, reference and excitation electrodes. The relative position of the fluid reservoir wall and the sensing electrode is also shown. D) Flex-PCB layout and images of fabricated sensor. CDC = capacitance-to-digital converter; CAPDAC = programmable on-chip digital-to-capacitance converter; GND = ground supply; PCB = printed circuit board; SDA/SCL = I²C communication pins; Vcc = power supply.

electrodes (diagonal pattern in Fig. 3A and B) are positioned around two sensing electrodes (dotted pattern) to measure fringing capacitance in the direction of the fluid as shown in Fig. 3C. We refer to this electrode design as excitation-sensing-excitation/inter-digitated arrangement (ESE-ID).

Two symmetrical gaps separate the excitation electrodes from the central sensing electrodes. The gap length is a parameter that affects the penetration depth of the most sensitive fringing pathways in other types of capacitive sensors using coplanar electrode arrangements.³⁸ Thus, L_{gap} can be iteratively adjusted to achieve optimal sensing in other applications with different fluid-sensor wall thicknesses. For our hydrostatic chamber prototype in Fig. 3A, $L_{\text{gap}} = 0.75$ mm was heuristically determined (from several design iterations) as sufficiently small to allow for capacitive sensing using our electrode geometry.

The AD7746 chip is connected to the front sensing electrode *via* the $C_{\text{sense}}(+)$ terminal, while the back sensing electrode is connected to the reference terminal $C_{\text{ref}}(-)$ to allow for differential measurements. The excitation electrodes of both sensing and reference planes are connected to the same excitation (EXC) terminal. The differential mode of the AD7746 chip was selected to maximize the robustness of capacitance measurements, while the addition of a symmetri-

cal reference arrangement at the back of the sensing layout was designed to maximize signal-to-noise ratio. This effect can be explained by referring to the expected fringing capacitance pathways on the presented mirrored electrode design (Fig. 3A), which shows a self-shielding effect within the sensor.

In commonly used capacitive sensing instruments, the region separating the electrodes from a target fluid is usually made of materials with low electrical permittivity (ϵ) such as plastic, glass or air.³⁹ This renders the capacitance due to plastic structures and air small as long as $\epsilon_{\text{Air}} < \epsilon_{\text{Plastic}} \ll \epsilon_{\text{Fluid}}$ is maintained. Thus, in our system the capacitance attributable to the region next to the fluid can be approximated to the total capacitance detected at the $C_{\text{sense}}(+)$ terminal. Furthermore, in previously described coplanar sensor designs,³⁹ the reference electrodes are usually situated in the same plane as the main sensing electrode (next to a region constantly filled with fluid). This traditional configuration simplifies the sensor compensation for different kinds of fluids and temperature changes, but it also brings several limitations in terms of footprint, minimum detectable fluid volume and achievable signal-to-noise ratio.

Our sensor prototype is implemented using a 3-layer flexible printed circuit board (Flex-PCB) with total thickness of 0.2 mm. Electrodes are defined as 0.5 oz copper layers with 17.5 μm thickness, while a 55 μm polyimide film was used as dielectric. A layer of dielectric film lies between the mirrored electrodes, as well as at the top and bottom of the sensor to protect the conductive material from corrosion (Fig. 3D). Connection traces between sensing circuit and electrodes use 0.127 mm traces. The electrodes are made with an original length $L_{\text{total}} = 20$ cm, and then are cut at the top end to fit the 6 cm fluid reservoir. This design allows for use of the sensor in longer hydrostatic chambers with minimum modification of the sensor design. The separation of the mirrored reference electrode from the main sensing plane was achieved using a 75 μm adhesive-polyimide-adhesive dielectric layer. Since increasing the distance between the central electrode and the fluid is expected to reduce sensitivity of the capacitive sensor, we aimed to reduce this separation considering the minimal resin thickness achievable with our 3D printing equipment without producing significant wall defects (~ 0.45 mm). This structural separation, in addition to the sensor's outer polyimide dielectric layer (55 μm), and the 1 mL syringe's wall thickness added to a minimum electrode-fluid separation for our testing setup of approximately 2 mm (shown in Fig. 3C). As with other capacitive sensing systems, the materials separating the electrodes from the fluid (*e.g.* syringe wall) are not expected to significant effects on the sensor's performance as long as the electrical permittivity of these materials are at least one order of magnitude less than that of the used fluid ($\epsilon_{\text{Plastic}} \ll \epsilon_{\text{Fluid}}$). Many other potential sources of variability affecting the performance of our capacitive sensor, such as electrode positioning, temperature changes, defects in structural materials and properties of the handled fluid, can all be automatically accounted for



using an initial calibration routine and by performing differential measurements.

The charge distribution imposed by this design directs the fringing fields of the sensing electrode plane preferably towards the fluid, while directing the fields of the reference electrode plane towards the back of the sensor. This configuration is more compact than traditional coplanar capacitive sensors, and enables the detection of smaller increases in fluid height and more effective compensation of external parasitic capacitances (*e.g.* user's movement). For this configuration, the capacitance associated with the sensing and reference coplanar electrodes can be approximated as:⁴⁰

$$C_{\text{sense}}(+)\approx\frac{\pi^*\varepsilon_0^*\varepsilon_s}{\ln\left(\frac{\pi(d-w)}{w+t_c}+1\right)}H_{\text{Fluid}}\quad(1)$$

$$C_{\text{ref}}(-)\approx\frac{\pi^*\varepsilon_0^*\varepsilon_R}{\ln\left(\frac{\pi(d-w)}{w+t_c}+1\right)}L_{\text{sensor}}\quad(2)$$

where $C_{\text{sense}}(+)$ and $C_{\text{ref}}(-)$ are the capacitances of the sensing and reference electrodes respectively, d is the average diameter of the fringing arcs between the sensing and the excitation electrodes. L_{sensor} is the total length of the sensor, H_{Fluid} is the fluid height, t_c is the thickness of the electrode conductor (assumed to be constant across all electrodes), and ε_0 is the permittivity of free space ($\varepsilon_0 \approx 8.9 \times 10^{-12}$ F m⁻¹). Due to symmetry, eqn (1) and (2) are only valid for the case in which the width of the sensing electrode (w) is approximately equal to the added width of both excitation electrodes. The relative permittivity associated with the sensing region (ε_s) and the reference region (ε_R) in eqn (1) and (2) can be determined by examining the ratio between the average diameter of fringing arcs (d) and the thickness of the material separating the fluid from the conductor (t_w).⁴¹ The thickness t_w considers the flex-PCB dielectric and plastic wall and is only used to determine ε_s and ε_R according to the following rules: for $d/t_w \gg 1$, $\varepsilon_s = \varepsilon_R \approx 1$; whereas for $d/t_w \approx 1$, $\varepsilon_s = (1 + \varepsilon_{\text{fluid}})/2 \approx 40$ and $\varepsilon_R = (1 + \varepsilon_{\text{plastic}})/2 \approx 1$. All previous approximations assume a relative permittivity for the fluid ($\varepsilon_{\text{fluid}}$) around 80 at 20 °C under a 1 kHz excitation, while $\varepsilon_{\text{plastic}} \approx 2$ for a bulk plastic dielectric material. Since our testing setup has $d = 2.5$ mm and $t_w = 2$ mm, it follows that $d/t_w = 1.25 \approx 1$ confirming that the second ratio condition applies. Dividing $C_{\text{ref}}(-)$ from $C_{\text{sense}}(+)$ and reordering terms we can approximate the fluid level in the reservoir as follows:⁴¹

$$\Delta H_{\text{Fluid}} \approx \left(\frac{\varepsilon_R}{\varepsilon_s}\right) \frac{C_{\text{sense}}(+)}{C_{\text{ref}}(-)} L_{\text{sensor}} - H_{\text{offset}}\quad(3)$$

where $\varepsilon_R/\varepsilon_s = (1 + \varepsilon_{\text{plastic}})/(1 + \varepsilon_{\text{fluid}})$ is just the proportionality constant in eqn (3), which depends on the target fluid and the bulk dielectric material. This constant can be calculated

for compensation purposes by performing a single capacitance measurement at a known fluid height. Finally, an additional compensation offset (H_{offset}) may be used to correct for mechanical inaccuracy due to sensor placement and to adjust for absolute height according to a common height reference as seen in eqn (3). The design and fabrication files for the implemented sensor are available for download in the ESI.†

II.a.3 Electronic feedback-control hardware. After acquisition of capacitance readings by the AD7746 chip, the digitized measurements are transmitted to a microcontroller board HUIZZAH® ESP8266 Feather (Adafruit Industries, New York, USA) *via* a ZIF connector and I²C serial communication terminals to perform feedback control computations. This board was selected as our control hardware due to its low-cost (<\$20 USD), ease of programming (using Arduino® syntax), availability of open-source design files, power efficiency and integrated wireless capabilities. This microcontroller board was then connected to a stackable custom-made active pumping board, and a FeatherWing OLED I/O interface board (Adafruit Industries, New York, USA) with assembled push buttons and a screen for offline parameter visualization. The detail in Fig. 1A shows the assembled control hardware, while additional images of the individual components of the controller can be found in Fig. S2 from the ESI.†

II.a.4 Bidirectional supply pumps. In order to actively control fluid height within the hydrostatic fluid chamber, two Bartels Microtechnik mp6 piezoelectric pumps (Servoflo, Lexington, USA) are driven in a bidirectional configuration by the pumping board using two commercially available mp6-OEM driver circuits (Servoflo, Lexington, USA).⁴² The flow-rate of these piezoelectric pumps is controlled using a pulse-width modulated (PWM) signal generated by two independent output channels from the HUIZZAH® ESP8266 Feather. This type of pump was selected due to its small size, fast response, high dynamic range, and chemical inertness. Other active bidirectional pumps may be used as long as their response time is faster than the characteristic time constant of the fluidic plant to be controlled. All PCB fabrication files to reproduce the pump board are included in the ESI.†

II.a.5 Control and user-interface software. The AD7746 acquisition routine and a proportional-integral (PI) control law were implemented using the Arduino® in-system programmer (ISP) on the ESP8266 microcontroller unit. We also used the Arduino and Adafruit ESP8266 libraries to facilitate integration of this platform. Capacitance measurements were obtained every 10 milliseconds using a timer-driven interrupt. A sensor calibration routine was also implemented, so as to allow the user to record a base capacitance offset as well as the calculation of the fluid-dependent proportionality constant ($\varepsilon_R/\varepsilon_s$). Routines to follow constant and pre-programmed dynamic fluid height profiles were also implemented in non-volatile memory of the ESP8266 (EEPROM). Acquired data and system control parameters were transmitted *via* USB and wirelessly *via* Wi-Fi to a laptop and then converted to CSV format for analysis. The code used for the conducted experiments is included in the ESI.†



II.a.6 Microfluidic devices and fluidic circuit. Two simple microfluidic devices were used to test usability and validate functionality of our feedback-controlled gravity-driven setup. These microfluidic devices are: 1) A custom fabricated single channel polydimethylsiloxane (PDMS) chip, and 2) A commercially available chip for cell culture as described later. For most of our tests the output channel of the driven microfluidic device was connected to a dripping recirculation pathway that maintained the outlet of the microfluidic channel at atmospheric pressure as depicted in Fig. 1A. It is also possible to use two of these hydrostatic pressure chambers at both ends of the microfluidic chips to allow for bidirectional gravity-driven flow, as well as regulating the device average pressure.

II.a.7 Structural and other components. Structural elements such as the sensor/chamber holder, controller box and microfluidic device holder were fabricated from stereolithographic (SLA) resin using a Form 2 3D printer (Formlabs, Somerville, USA). The hydrostatic chamber and sensor holder are made from clear resin to allow for direct optical visualization of the monitored fluid front for validation purposes. The ESE-ID flexible capacitive sensor was positioned at the desired distance from the fluidic chamber using a guiding slot within the 3D printed structure, and was permanently fixed in this position using PDMS. The design and fabrication files of these structural components are also included in the ESI.†

II.b System testing and validation

II.b.1 Validation of capacitance measurements. In order to characterize the accuracy of the capacitance measurements obtained by our sensor, we first isolated the AD7746 circuit of the Flex-PCB. We then placed capacitors of known values between the sensor's $C_{\text{sense}}(+)$ and EXC terminals to record their values. The tested capacitances were verified using a calibrated E4981A capacitance meter (Keysight Technologies, Santa Rosa, USA) for comparison purposes. The results of this test are shown in Fig. 4A.

II.b.2 Basic fluid-height tracking. After characterization of the sensing circuit, a basic fluid height “challenge” within an isolated hydrostatic chamber with an embedded capacitive sensor was performed in triplicate using a 1× phosphate buffered saline (PBS) solution. The fluid level in this reservoir was controlled using a calibrated syringe pump 11-Pico PLUS Elite (Harvard Apparatus, Holliston, USA). This pump was programmed to produce a dynamic oscillating change in fluid volume over the reservoir's entire height range (6 cm) during a 20 s period. Readings from our capacitive sensor were taken every 10 ms (without averaging) and compared to the fluid volume supplied by the syringe pump. Fluid height in the chamber was also verified using video recordings and image processing to track the fluid-air interface visualized through the translucent regions of the fluid reservoir. The results are shown in Fig. 4B.

II.b.3 Gravity-driven flow rate validation. To validate flow rates established using a given hydrostatic pressure (*i.e.* fluid




Fig. 4 A) Results of capacitive sensing circuit validation experiments. Capacitance measurements with the developed sensing circuit matched real values within <0.1 pF error. Circles denote average of experiment conducted in triplicate at a single capacitance value with the error bars shown in green. A linear fit of the measured values is also shown. B) Results of basic fluid-level tracking challenge using capacitive sensing. The dashed profile (red) is the expected fluid-height as imposed by the calibrated syringe pump. Black points refer to the averaged capacitive sensor output for the three replicates at each time point. C) Measured and calculated flow rates as a function of total fluid height in gravity-driven pump. The measured values (circles) with green error bars, closely approximate the theoretical values calculated using Poiseuille's equation (asterisks). A linear fit of the measured values is also shown.

height), we connected our pump to a 50 cm long silicone tube with inner diameter of 0.79 mm ($1/32$ " and measured the mass of 1× PBS displaced over a one minute duration ($n = 3$) using a laboratory grade scale. The Poiseuille equation was used to calculate the expected flow rate given a hydrostatic pressure difference $\Delta P = 8 \mu LQ / (\pi R^4)$, where ΔP is the hydrostatic pressure difference, L is the length of the tubing, μ is the dynamic viscosity of PBS which is assumed to be close to that of water (8.9×10^{-4} Pa s at 25 °C), Q is the volumetric flow rate and R is the inner radius of the tubing. Additionally, the hydrostatic pressure was calculated as $\Delta P = \rho gh$, where ρ is the density of PBS which is approximately that of water (1 kg m^{-3}), g is gravity and h is the total fluid height above the outlet feeding to the collection tubes used for weight measurement. In order to achieve high flow rates, we



added an additional 53 mm offset to the fluid column by placing the collection tubes below the gravity driven pump outlet. The obtained flow-rate measurements (in triplicate) were compared to expected values using Poiseuille's equation as shown in Fig. 4C.

II.b.4 Characterization of fluid-type dependency. As described earlier, the gain of our capacitive sensor design is expected to change depending on the charge distribution resulting from the conductivity and the electrical permittivity of the target fluid. Therefore, changes in charged solute concentration can affect these readings significantly. To characterize such effect, we conducted a titration experiment in triplicates using 1 \times , 0.5 \times , 0.25 \times , 0.125 \times and 0.0625 \times PBS, with deionized (DI) water and a steel inserts of known lengths as controls. Two inserts were made by cutting 10 and 30 mm sections from a tight-tolerance multipurpose O1 tool steel rod with 0.1750" diameter (McMaster Carr, Robbinsville, USA). The associated capacitance value from all fluid conditions and materials was recorded at two fixed fluid heights: lower bound ($H_0 = 10$ mm) and upper bound ($H_1 = 30$ mm) as shown in Fig. 5A. To confirm experimental results, we also computed the electric displacement flux density (D) around relevant target fluids (*i.e.* 1 \times PBS solution, DI-water), *via* finite element analysis (FEA) using the AC conduction field solver of ANSYS Maxwell software (ANSYS, Inc., Canonsburg, USA) as shown in Fig. 5D and E. In these simulation, the two middle electrodes were fixed to zero voltage, and the four electrodes on the side were excited with a sinusoidal voltage of 5 V amplitude at an excitation frequency $f_{\text{EXC}} = 32$ kHz. These conditions are analogous to the settings used by the AD7746 24 bit Σ - Δ capacitance-to-digital converter in the developed sensor.

II.b.5 Flow-rate tracking and calibration of external pneumatic micropumps. Since accurate contactless assessment of flow in open-channel microfluidics is a valuable potential application for the present technology, we conducted a series of proof-of-concept experiments to demonstrate flow-rate extraction and open-loop pumping calibration based on continuous fluid-level monitoring. The same previously described fluid-height tracking methodology (at 37 °C and 95% humidity) was used, except that in these experiments, the calibrated syringe pump was programmed to impose specific flow-rate profiles (instead of volumetric changes) ranging from 100 nL min^{-1} to 1 mL min^{-1} . These conditions included constant, ramp and intermittent flow-rates in input and output mode. Fluid-height changes over time were measured in triplicates and compared to theoretical flow estimates using the pre-programmed flow-rate parameters imposed by the calibrated syringe pump. Extraction of flow-rate was approximated by generating a linear fit of fluid volume change over time every 1000 samples ($t = 10$ s). Experiments were limited to a maximum input/output volume ($V_{\text{max}} = 0.5$ mL) and a maximum experimental period ($T < 400$ s). The results are shown in Fig. S3 of the ESI.†

In a subsequent experiment, this height-based flow-rate tracking methodology was used to characterize and calibrate




Fig. 5 A) Results on dependence of capacitive sensing on fluid type. Use of fluids with high concentration of solutes forming free ions lead to higher capacitance measurements and ranges. Error bars representing standard deviation of upper measurements ($h = 30$ mm) are shown in orange, while error bars of the lower measurements ($h = 10$ mm) are shown in green. B) Capacitive fringing model for electrolytic fluids with free ions. C) Capacitive fringing model for non-conductive fluids with high dielectric constant. Fringes at the fluid-wall interface exhibit a change in angle towards the symmetry axis (not shown) depending on properties of wall material and fluid. D) Finite element analysis (FEA) of the electric displacement flux density for 1 \times PBS target fluid. E) FEA simulation of electric flux density for DI-water as target. In both D) and C) the blue circle is the target liquid, and the outer box is plastic material ($\epsilon_r = 2.7$, $\sigma = 0$ S m^{-1}) following the same dimensions as the physical prototype chamber. Outside the plastic box region there is air. The thin orange contour surrounding the sensor electrodes is a polyamide film ($\epsilon_r = 4.3$, $\sigma = 0$ S m^{-1}). The color of the vectors represents the magnitude of the flux density D (nC m^{-2}) shown in the middle color bar. The size of the arrows is proportional to the magnitude of the flux density to facilitate interpretation.

two presumably identical open-loop pneumatic diaphragm micropumps. These micropumps were set to supply and extract fluid from the same monitored reservoir at 1 $\mu\text{L s}^{-1}$. These pumps were fabricated in acrylic using a CNC mill and a thin polyurethane membrane according to a previously reported protocol.⁴³ Both pumps were actuated for 40 min at 1 Hz (stroke volume = 1 μL) under 37 °C and 95% humidity



to observe fluid-height drift (in triplicate). After confirming adequate operation of both pumps, any observed volume drift in the monitored reservoir is assumed to be caused by small differences in input/output pumping performance attributable to fabrication variations or head pressure effects. After combined drift characterization, flow-rates were independently tracked in triplicate for each pump to recalibrate their actuation frequency and adjust for errors. After calibration, fluid-height drift was again characterized for 40 min and compared against uncalibrated behavior ($n = 3$) as shown in Fig. 6.

II.b.6 Validation of closed-loop control of gravity driven pump. The performance, robustness and dynamic range of the proposed closed-loop feedback control system was characterized using the entire monitored setup as illustrated schematically in Fig. 1A. A single-channel microfluidic device made in a PDMS (length = 50 mm, width = 1 mm and height = 0.2 mm) was connected to the outlet of the hydrostatic chamber with capacitive sensing. After the chip was connected to the setup, sensor calibration and flow testing was performed to assess the emptying time constant of the hydrostatic chamber given the fluidic resistance from the connected microfluidic chip. After this, a closed-loop feedback control mode was activated to automatically control fluid input/output from the secondary piezoelectric pumps to maintain a target height ($\Delta H = 30$ mm). For this experiment, the hydrostatic chamber was filled with $1\times$ PBS, and was monitored over a period of 48 hours inside an incubator (at

37°C , 5% CO_2 and 95% humidity). Presence of flow through the microfluidic device was confirmed by observing dripping fluid within the secondary recirculation container. This experiment was conducted in triplicate and the fluid height was also verified using video recordings of the fluid front to assess drift as shown in Fig. 7A.

After verifying adequate closed-loop performance for a constant fluid-level set-point, a dynamic fluid-level target experiment was conducted under similar experimental conditions. However, in this case the target was pre-programmed to be a dynamic fluid-level profile stored in non-volatile memory of the MCU. This profile was a 40 min sequence including constant, sine, triangular, saw tooth and step waveforms (two periods each). This experiment was also conducted in triplicate within an incubator (37°C at 95% humidity) using video recordings to verify location of the fluid front. Results are shown in Fig. 7A.

II.b.7 Cell culture experiments. In order to show biocompatibility of the setup, a 24 hour cell culture experiment was performed using iPSC-derived vascular endothelial cells as an example cell type. In this test, an IBIDI μ -Slide VI 0.4 channel slide (IBIDI, Martinsried, Germany) was coated with human fibronectin (Life Technologies, Woburn, USA) at a concentration of $30\ \mu\text{g mL}^{-1}$ for 1 hour at room temperature. Induced Pluripotent Stem Cell (iPSC) derived endothelial cells (CDI, Madison, WI) were seeded in all the channels, allowed to adhere for 3 hours then excess cells were washed away using two successive media changes. Results are shown in Fig. 8. The device was cultured for 24 hours in the supplier-recommended media under standard incubator conditions.




Fig. 6 A) Testing block with six parallel pneumatic diaphragm micropumps. Two pumps were used to drive fluid in and out the monitored fluid reservoir. B) Fluid-height change over time (40 min) as measured by the capacitive sensor in the system driven by unbalanced input/output micropumps. C) Overlay of isolated experimental input and output behavior of each pneumatic diaphragm micropump. Blue profile denotes fluid-height over time produced by pump A (input) alone, while red profile shows the same for pump B (output). D) Fluid-height change over time (40 min) in the same system after micropump calibration using extracted flow-rate from capacitive sensing readout. Black points denote average for the three replicates at each time point. Standard deviation is shown in light green.




Fig. 7 A) Results of closed-loop feedback control for a constant set point ($\Delta H = 30$ mm). Black points refer to the averaged capacitive sensor output for the three replicates at each time point. Error bars are shown in green. B) Results of closed-loop feedback control for a dynamic set point using constant, sine, triangular, saw tooth and step waveforms. Black points refer to the averaged capacitive sensor output for the three replicates at each time point. Error bars representing standard deviation are shown in green.






Fig. 8 A) Closed-loop gravity-driven setup used for cell-culture experiment. B) Confocal microscopy of endothelial cells cultured under static conditions as control. C) Confocal microscopy of endothelial cells cultured under 4 cm of closed-loop gravity-driven flow. Both cultures present with no observable differences, which suggests appropriate biocompatibility of the system.

After assembly, the gravity-driven setup was sterilized by circulating 70% ethanol through the entire fluidic circuit for 5 min. Ethanol was removed by air drying within a sterile hood and then flushing with two cycles of sterile DI water. After sterilization, the seeded IBIDI μ -Slide VI microfluidic channels were connected to the gravity-driven setup within a sterile hood. The hydrostatic chamber was then programmed to maintain a constant 40 mm fluid-height (392 Pa inlet pressure) to drive flow through the chip which had its outlet at atmospheric pressure through the dripping recirculation circuit. Laminar flow was assumed given the imposed pressure gradient and the rectangular design of the IBIDI device channel (length = 17 mm, width = 3.8 mm and height = 0.4 mm). The Poiseuille equation was used to calculate the expected flow rate given a hydrostatic pressure difference $Q = \Delta P/R$; where Q is the volumetric flow rate, ΔP is the hydrostatic pressure difference (392 Pa for a 40 mm media height) and R is the total resistance of the fluidic circuit ($R_{\text{channel}} + R_{\text{tubing}}$). For the IBIDI channel, which has a rectangular cross section, the fluidic resistance can be calculated as:^{44–46}

$$R_{\text{channel}} = \frac{12\mu L}{wh^3 \left(1 - 0.63 \frac{h}{w}\right)} = 6.21 \times 10^8 \text{ Pa s m}^{-3}$$

where μ is the dynamic viscosity of media which was assumed to be close to that of water (6.94×10^{-4} Pa s at 37 °C), L is the length of the channel, w is the width and h is the

height. For the tubing, which has a circular cross-sectional area with inner diameter 1/32 inch and total length of 20 cm, the fluidic resistance can be calculated as:⁴⁵

$$R_{\text{tubing}} = \frac{8\mu L}{\pi r^4} = 1.48 \times 10^{10} \text{ Pa s m}^{-3}$$

where r is the inner radius of the tubing, L is the length of the tubing, μ is the dynamic viscosity of media which was assumed to be close to that of water (6.94×10^{-4} Pa s at 37 °C). Using this information, we calculated the flow rate Q to be $2.64 \times 10^{-8} \text{ m}^3 \text{ s}^{-1}$ or 1.59 mL min^{-1} at this specific height. The wall shear stress, τ , imposed on the endothelial cells under this flow rate can be calculated as:⁴⁷

$$\tau = \frac{6\mu Q}{h^2 w} = 1.8 \text{ dyne per cm}^2$$

here τ is in dyne per cm^2 , μ is in poise, Q is in $\text{cm}^3 \text{ s}^{-1}$ and h and w are in cm and are the height and width of the IBIDI channel respectively. After 24 hours of culture, cells were fixed with 4% PFA, stained with DAPI and rhodamine and imaged under an EVOS inverted microscope (Life Technologies, Woburn MA, USA) with a 20 \times objective.

II.c Statistical analysis

All validation experiments were conducted in triplicates. Error bars represent standard deviation in all figures. Calculations and plots were generated with Graphpad Prism 7 (GraphPad Software Inc.; La Jolla, USA).

III. Results & discussion

III.a Accurate capacitance fluid sensing and fluid tracking

The results of the initial characterization experiments used to assess the accuracy of capacitance readings and basic fluid level tracking capabilities of our sensing circuit are shown in Fig. 4. Measured values of known capacitors directly connected to the $C_{\text{sense}(+)}$ and C_{exc} terminals in our sensing circuit appear to match those of the E4981A capacitance meter within <0.1 pF error (Fig. 4A). Readings from the capacitive sensing circuit also appear to be linearly related to those obtained from the calibrated reference meter suggesting appropriate implementation of the sensor at the PCB level. Fig. 4B shows the aggregated results for three basic fluid-level tracking challenges using the fluid chamber with our proposed excitation-sensing-excitation inter-digitating capacitive sensor arrangement (ESE-ID) as depicted in Fig. 2C. The dashed red profile in Fig. 4B is the known fluid-height as imposed by the calibrated syringe pump. Black data points refer to the averaged capacitive sensor output for the three replicates at each time point. The average standard deviation across all samples in this experiment was <250 μm when compared to the fluid height imposed by the calibrated syringe pump. Fig. 4C shows both measured and estimated flow-rates in our gravity-driven pump as a function of set



fluid height. While experimental measurements appear to closely follow the theoretical values predicted by Poiseuille's equation, there also seems to be a slight overestimation of flow rate near the most upper level of the capacitive electrodes. While not fully explored in this work, such effect appears to be caused by an unmolded fringing capacitance at this boundary, which may be further corrected *via* software using a linear calibration curve. Fig. 4C shows an apparent minimum fluid height of 54 mm for this specific test; nonetheless, this included a 53 mm offset from the measuring outlet as described in section II.b.3. Considering this offset, it can be recognized that the actual minimum fluid height controlled for this test corresponded to approximately 1 mm. Despite the fact that our capacitive sensor was able to detect complete emptying of the fluid chamber, we selected this minimum fluid level in order to avoid partial emptying or significant meniscus effects such as edge pinning during this test. Error bars are shown in green for all measurements in Fig. 4. If further error reduction is required, averaging techniques and low-pass filtering may also be implemented in longer experimental time scales.

III.b Capacitance readings depend on fluid conductivity

Fig. 5A shows the results from the titration experiments conducted to characterize performance of our capacitive sensor based on fluid conductivity. Capacitance readings for upper ($H_1 = 30$ mm) and lower ($H_0 = 10$ mm) levels were smallest and closest together for deionized (DI) water and increased as electrolyte concentration increased to $1\times$ PBS (conductivity @25 °C = 1.6 S m^{-1}). Small steel rods of both lengths were also placed inside the monitored reservoir to assess signal in the presence of a known perfect conductor. Higher dilutions ($0.5\times$ to $0.0625\times$ PBS) led to a range reduction and lower absolute capacitance, reaching a minimum in DI-water lacking solutes (conductivity @25 °C = $0.055\text{ }\mu\text{S cm}^{-1}$).

This behavior can be explained by the model shown in Fig. 5B and C. In the case of water with dissolved solutes forming free ions (Fig. 5B), a large number of sensing fringing paths become terminated near the fluid-reservoir boundary leading to higher capacitance measurements. Conversely, in the case of pure water (Fig. 5C), fringing capacitance takes longer uninterrupted paths through the bulk of the fluid, generating smaller capacitance measurements. Fig. 5D shows the simulated displacement flux density around the $1\times$ PBS solution ($\epsilon_{r,\text{PBS}} = 80$, $\sigma_{\text{PBS}} = 1.45\text{ S m}^{-1}$). The charge relaxation time of the $1\times$ PBS solution is $\tau_{\text{PBS}} = \epsilon_{r,\text{PBS}} \epsilon_0 / \sigma_{\text{PBS}} = 488\text{ ps}$, which corresponds to a break frequency $f_{\text{PBS}} = 2.04\text{ GHz}$. This break frequency is orders of magnitude higher than the 32 kHz sensor excitation frequency used for the sensor.

Since $f_{\text{PBS}} \gg f_{\text{EXT}}$, the $1\times$ PBS solution behaves like a conductor, accumulating free charges on its surface and increasing the apparent capacitance seen from the sensor electrodes. The induced surface charges terminate the flux density vectors, and therefore the flux density vectors are invisible inside the $1\times$ PBS solution in Fig. 5D. For the case of

DI-water ($\epsilon_{r,\text{DI}} = 80$, $\sigma_{\text{DI}} = 5.5\text{ }\mu\text{S m}^{-1}$), Fig. 5C shows the field lines schematically, while Fig. 5E shows the simulated flux density field. The charge relaxation time of the DI-water is $\tau_{\text{DI}} = \epsilon_{r,\text{DI}} \epsilon_0 / \sigma_{\text{DI}} = 129\text{ }\mu\text{s}$, which corresponds to a break frequency $f_{\text{DI}} = 7.76\text{ kHz}$. Since $f_{\text{DI}} < f_{\text{EXT}}$, the DI-water behaves like an insulator, which explains why the flux density vectors induced inside the DI-water follows the pattern shown in Fig. 5C and E. These results show that calibration is required to adjust for potential differences in fluid conductivity. Further sections of this work make use of sensors calibrated for $1\times$ PBS and culture media.

III.c Accurate flow-rate tracking and calibration of external pneumatic micropumps

Fig. S3 of the ESI† shows the results of an extended series of experiments carried out to verify if accurate flow-rate calculations were achievable from analyzing fluid-height changes over time using our capacitive sensor. From these experiments, it was confirmed that a variety of flow-rate conditions can be inferred based on this sensor's output and that these values correspond to the programmed settings in the calibrated syringe pump used to impose flow. Fig. 6 shows the results of the characterization and calibration experiment of two open-loop pneumatic diaphragm micropumps (Fig. 6A) feeding and extracting fluid from the same monitored reservoir at a nominal rate of $1\text{ }\mu\text{L s}^{-1}$. Over the course of 40 min, a 0.2 mL decrease in fluid volume was observed in the monitored reservoir (Fig. 6B), showing that the output pump flow-rate was slightly greater than the input pump flow rate despite being actuated at the same frequency.

Independent flow-rate characterization of each of these micropumps using our capacitive sensor revealed an 8.3% mismatch between the output and input micropump flow rates. This difference explains the observed decreasing height drift and is well within the 10–15% expected error usually reported for these type of pumping systems.²⁴ After recalibration of the actuation frequency for both micropumps based on these readings, fluid-height drift appeared to be corrected for the same 40 min period as compared to the uncalibrated behavior. In general, long-term drift is expected to be present for a wide range of open-loop micropumps, due to back pressure, use-induced stress or solute deposition may sporadically change stroke volume. Thus, our sensor could be a valuable addition to open-channel microfluidic systems requiring accurate flow control, or as a new way to assess flow-rates and perform pump calibrations on demand.

III.d Reliable closed-loop control of gravity driven pump

From available approaches for fluid handling in micro- and mesofluidic devices, gravity-driven systems have historically been considered among the most robust, simple and convenient to use.¹³ Reservoirs acting as gravity-driven pumps are relatively low-cost, can achieve a wide range of flow-rates, rarely lead to bubble stagnation and usually do not require external power to impose flow.^{13,48} However, traditional



gravity-driven pumps (with vertically positioned reservoirs) can only produce unidirectional transient flows as the liquid level in the reservoir decreases.^{13,16,22} This situation leads to a time-dependent reduction in achievable flow rate proportional to the decline in hydrostatic pressure. This transient mode of operation is a key limitation of most gravity-driven systems, especially in long-term cell culture applications.²²

Recent modifications of gravity-driven systems have been reported to provide nearly constant flow rates either through the use of horizontal reservoirs (setting a deterministic internal fluid height) or through the use of a large vertical reservoir (maintaining fluid heights nearly constant during limited operation times).^{49–51} Despite the advantages that these modifications may provide, most currently reported gravity-driven microfluidic devices still remain open-loop in nature, are cumbersome to continuously monitor and cannot deliver bidirectional, smooth, reconfigurable flow over long periods of time. Without closed-loop feedback control, the emptying time for an initial 30 mm fluid column at the hydrostatic chamber in connection with the used single-channel microfluidic chip was approximately 265 s.

Fig. 7 shows the typical closed-loop response provided by our gravity-driven microfluidic setup to a static (Fig. 7A) and dynamic (Fig. 7B) target. Both graphs show the aggregated height measurements for three experimental replicates as a function of time. Capacitive readouts were acquired every 10 ms, but were continuously averaged over 10 samples ($T = 100$ ms) in both cases to feed the control algorithm. The maximum standard deviation across the entire 48 h testing period for the constant target set point ($\Delta H = 30$ mm) was 0.65 mm, with a calculated drift of $<1\%$. Fig. 7B, shows the results obtained over a 40 min period using a dynamic pre-programmed set point.

The constant, sine, triangular, saw tooth and step waveforms were all recognizable and accurately followed with less than 5% error. Overshooting decaying oscillations were observable at the high-frequency transitions for both step-like cycles, which is characteristic of many second order systems using closed-loop feedback control as it reflects the control loop dynamics. In our augmented gravity-driven setup, fluid height appears to be a useful target variable allowing us to accurately control both pressure and flow-rate in this system. While these variables are commonly measured and used to control large-scale gravity-dominated fluidic systems (volumes > 10 L),^{52,53} this is the first time, to our knowledge, that capacitive sensing has been adapted to control such parameters in an open-well microfluidic system. Thus, we believe that previous limitations associated with sensing accuracy, miniaturization, ease of use and cost of implementation, may be addressed by our proposed sensor design.

III.e Biocompatibility for continuous cell culture

Finally, Fig. 8A shows the assembled and sterilized gravity driven setup used for our cell culture proof-of-concept. Optical analysis of both static (Fig. 8B) and gravity-driven

(Fig. 8C) cultures confirmed viability. No differences in endothelial growth was observed in the test samples compared to the control, which suggests adequate biocompatibility of the designed fluidic circuit, as expected due to the use of inert materials in the system's fluidic pathways. Given the low shear stress we exposed the cells to (1.8 dyne per cm^2), we did not expect alignment in the direction of flow as such response for endothelial cells typically requires a shear >10 dyne per cm^2 and in previous characterization of iPSC-derived endothelial cells a 20 dyne per cm^2 shear stress was used.⁵⁴

Conclusions

Previous to this work, accurate measurement of fluid-height in open reservoirs for microfluidic devices had been considered a challenging task, especially in applications using total volumes below 1 mL. To our knowledge we are the first group to use this type of sensing technology to perform continuous measurements of fluid height changes in open micro- and meso-fluidic reservoirs at submillimeter spatial resolution and millisecond temporal resolution to provide closed-loop control of pressure gradients and flow through a gravity driven system.

Though limited in some ways (*e.g.* dependent on fluid properties and requiring non-conducting fluid chambers) we believe the design principles presented here can be used to adapt such sensors to a wide range of open-channel microfluidic devices. In particular, we believe this technology may allow for continuous fluid height monitoring in high throughput microtiter plates and cell culture platforms with the help of robotic liquid handlers, or even in the monitoring of gravity driven chambers to measure permeability of small porous materials such as scaffolds and hydrogels.

Conflicts of interest

The authors confirm that there are no conflicts to declare.

Acknowledgements

The authors would like to thank the microfluidics community who have provided both direct and indirect input for the design specifications the presented sensor and system. This work was funded under the Defense Advanced Projects Agency (DARPA) Microphysiological Systems Program (W911NF-12-2-0039).

References

- 1 S. K. Vashist, P. B. Lippa, L. Y. Yeo, A. Ozcan and J. H. Luong, Emerging technologies for next-generation point-of-care testing, *Trends Biotechnol.*, 2015, 33(11), 692–705.
- 2 C. K. Byun, K. Abi-Samra, Y. K. Cho and S. Takayama, Pumps for microfluidic cell culture, *Electrophoresis*, 2014, 35(2–3), 245–257.
- 3 L. Chen, S. Lee, J. Choo and E. K. Lee, Continuous dynamic flow micropumps for microfluid manipulation, *J. Micromech. Microeng.*, 2007, 18(1), 013001.



- 4 D. K. Abeywardana, A. P. Hu and Z. Salcić, Pulse Controlled Microfluidic Actuators with Ultra Low Energy Consumption, *Sens. Actuators, A*, 2017, 263, 8–22.
- 5 A. T. Al-Halhouli, M. I. Kilani and S. Büttgenbach, Development of a novel electromagnetic pump for biomedical applications, *Sens. Actuators, A*, 2010, 162(2), 172–176.
- 6 C. H. Ahn and M. G. Allen, Fluid micropumps based on rotary magnetic actuators, in *Proceedings of MEMS*, 1995, vol. 95, pp. 408–412.
- 7 X. Huang and T. K. Chuan, MEMS-micropumps: a review, in *Journal of Fluids Engineering-Transactions of the ASME*, 2002.
- 8 T. Satoh, G. Narazaki, R. Sugita, H. Kobayashi, S. Sugiura and T. Kanamori, A pneumatic pressure-driven multi-throughput microfluidic circulation culture system, *Lab Chip*, 2016, 16(12), 2339–2348.
- 9 J. Gross, Chemically driven liquid delivery pumping device, US Pat. 5871125, Elan Medical Technologies Limited, 1999.
- 10 Z. R. Xu, C. G. Yang, C. H. Liu, Z. Zhou, J. Fang and J. H. Wang, An osmotic micro-pump integrated on a microfluidic chip for perfusion cell culture, *Talanta*, 2010, 80(3), 1088–1093.
- 11 B. D. Iverson and S. V. Garimella, Recent advances in microscale pumping technologies: a review and evaluation, *Microfluid. Nanofluid.*, 2008, 5(2), 145–174.
- 12 B. Weigl, G. Domingo, P. LaBarre and J. Gerlach, Towards non-and minimally instrumented, microfluidics-based diagnostic devices, *Lab Chip*, 2008, 8(12), 1999–2014.
- 13 J. Y. Chang, S. Wang, J. S. Allen, S. H. Lee, S. T. Chang, Y. K. Choi, C. Friedrich and C. K. Choi, A novel miniature dynamic microfluidic cell culture platform using electro-osmosis diode pumping, *Biomicrofluidics*, 2014, 8(4), 044116.
- 14 W. C. Nelson and C. J. C. Kim, Droplet actuation by electrowetting-on-dielectric (EWOD): A review, *J. Adhes. Sci. Technol.*, 2012, 26(12–17), 1747–1771.
- 15 D. R. Hewitt, J. A. Neufeld and N. J. Balmforth, Shallow, gravity-driven flow in a poro-elastic layer, *J. Fluid Mech.*, 2015, 778, 335–360.
- 16 M. Marimuthu and S. Kim, Pumpless steady-flow microfluidic chip for cell culture, *Anal. Biochem.*, 2013, 437(2), 161–163.
- 17 T. Kim and Y. H. Cho, A pumpless cell culture chip with the constant medium perfusion-rate maintained by balanced droplet dispensing, *Lab Chip*, 2011, 11(10), 1825–1830.
- 18 N. Nam-Trung, H. Xiaoyang and T. K. Chuan, Mems-micropumps: a review, *Trans. ASME*, 2002, 124, 384–392.
- 19 C. Y. Lee, H. T. Chang and C. Y. Wen, A MEMS-based valveless impedance pump utilizing electromagnetic actuation, *J. Micromech. Microeng.*, 2008, 18(3), 035–044.
- 20 V. Tandon, W. S. Kang, T. A. Robbins, A. J. Spencer, E. S. Kim, M. J. McKenna, S. G. Kujawa, J. Fiering, E. E. Pararas, M. J. Mescher and W. F. Sewell, Microfabricated reciprocating micropump for intracochlear drug delivery with integrated drug/fluid storage and electronically controlled dosing, *Lab Chip*, 2016, 16(5), 829–846.
- 21 C. Yamahata, C. Lotto, E. Al-Assaf and M. A. M. Gijjs, A PMMA valveless micropump using electromagnetic actuation, *Microfluid. Nanofluid.*, 2005, 1(3), 197–207.
- 22 A. J. Mäki, S. Hemmilä, J. Hirvonen, N. N. Girish, J. Kreutzer, J. Hyttinen and P. Kallio, Modeling and experimental characterization of pressure drop in gravity-driven microfluidic systems, *J. Fluids Eng.*, 2015, 137(2), 021105.
- 23 R. Gorkin, J. Park, J. Siegrist, M. Amasia, B. S. Lee, J. M. Park, J. Kim, H. Kim, M. Madou and Y. K. Cho, Centrifugal microfluidics for biomedical applications, *Lab Chip*, 2010, 10(14), 1758–1773.
- 24 K. Domansky, W. Inman, J. Serdy, A. Dash, M. H. Lim and L. G. Griffith, Perfused multiwell plate for 3D liver tissue engineering, *Lab Chip*, 2010, 10(1), 51–58.
- 25 C. Y. Chan, P. H. Huang, F. Guo, X. Ding, V. Kapur, J. D. Mai, P. K. Yuen and T. J. Huang, Accelerating drug discovery via organs-on-chips, *Lab Chip*, 2013, 13(24), 4697–4710.
- 26 Y. S. Heo, L. M. Cabrera, J. W. Song, N. Futai, Y. C. Tung, G. D. Smith and S. Takayama, Characterization and resolution of evaporation-mediated osmolality shifts that constrain microfluidic cell culture in poly (dimethylsiloxane) devices, *Anal. Chem.*, 2007, 79(3), 1126–1134.
- 27 J. F. Lo, E. Sinkala and D. T. Eddington, Oxygen gradients for open well cellular cultures via microfluidic substrates, *Lab Chip*, 2010, 10(18), 2394–2401.
- 28 E. Ghafar-Zadeh, M. Sawan, E. Ghafar-Zadeh and M. Sawan, *CMOS capacitive sensors for lab-on-chip applications*, Springer, 2010, pp. 25–34.
- 29 D. C. Stillinger, M. J. Head, K. N. Helland and C. W. Van Atta, A closed-loop gravity-driven water channel for density-stratified shear flows, *J. Fluid Mech.*, 1983, 131, 73–89.
- 30 N. T. Nguyen, Micromachined flow sensors—a review, *Flow Meas. Instrum.*, 1997, 8(1), 7–16.
- 31 B. Kuczenski, P. R. LeDuc and W. C. Messner, Pressure-driven spatiotemporal control of the laminar flow interface in a microfluidic network, *Lab Chip*, 2007, 7(5), 647–649.
- 32 M. F. A. Rahman, A. A. Manaf and M. R. Arshad, *Capacitive effect of coplanar electrodes partially outside the microchannel region for underwater microfluidic-based sensor*, 2013.
- 33 P. Wells, Capacitive fluid level sensor, US Pat. 5042299, IIMorrow, Inc., 1991.
- 34 C. Elbuken, T. Glawdel, D. Chan and C. L. Ren, Detection of microdroplet size and speed using capacitive sensors, *Sens. Actuators, A*, 2011, 171(2), 55–62.
- 35 J. Shih, J. Xie and Y. C. Tai, Surface micromachined and integrated capacitive sensors for microfluidic applications, in *TRANSDUCERS, Solid-State Sensors, Actuators and Microsystems, 12th International Conference*, IEEE, 2003, vol. 1, pp. 388–391.
- 36 E. Ghafar-Zadeh and M. Sawan, A hybrid microfluidic/CMOS capacitive sensor dedicated to lab-on-chip applications, *IEEE Trans. Biomed. Circuits Syst.*, 2007, 1(4), 270–277.
- 37 Z. W. Wei, M. Qin and Q. A. Huang, A Novel 2-D Capacitive Silicon Flow Sensor, in *Sensors, 2007 IEEE*, IEEE, 2007, pp. 888–891.



- 38 X. B. Li, S. D. Larson, A. S. Zyuzin and A. V. Mamishev, Design principles for multichannel fringing electric field sensors, *IEEE Sens. J.*, 2006, 6(2), 434–440.
- 39 C. S. Walker, *Capacitance, inductance, and crosstalk analysis*, Artech House, 1990.
- 40 J. Wang, Liquid level sensing using capacitive-to-digital converters, *Analog Dialogue*, 2015, 49(4), 1–3.
- 41 C. S. Walker, *Capacitance, Inductance and Crosstalk Analysis*, Artech House, 1990, ISBN: 978-0890063927.
- 42 T. Kassis, P. M. Perez, C. J. W. Yang, L. R. Soenksen, D. L. Trumper and L. G. Griffith, MICCS: A Fully Programmable Multipurpose Integrated Cell Culture System, *bioRxiv* 192047, DOI: 10.1101/192047.
- 43 W. Inman, K. Domansky, J. Serdy, B. Owens, D. Trumper and L. G. Griffith, Design, modeling and fabrication of a constant flow pneumatic micropump, *J. Micromech. Microeng.*, 2007, 17(5), 891.
- 44 L. A. Godwin, K. S. Deal, L. D. Hoepfner, L. A. Jackson and C. J. Easley, Measurement of microchannel fluidic resistance with a standard voltage meter, *Anal. Chim. Acta*, 2013, 758, 101–107.
- 45 S. Attiya, et al., Design of an interface to allow microfluidic electrophoresis chips to drink from the fire hose of the external environment, *Electrophoresis*, 2001, 22, 318–327.
- 46 D. Kim, N. C. Chesler and D. J. Beebe, A method for dynamic system characterization using hydraulic series resistance, *Lab Chip*, 2006, 6, 639.
- 47 H. J. Kim, D. Huh, G. Hamilton and D. E. Ingber, Human gut-on-a-chip inhabited by microbial flora that experiences intestinal peristalsis-like motions and flow, *Lab Chip*, 2012, 12, 2165–2174.
- 48 P. Buchling and S. Kandlikar, Enhanced flow boiling of ethanol in open microchannels with tapered manifolds in a gravity-driven flow, *J. Heat Transfer*, 2016, 138(3), 031503.
- 49 S. Takayama, J. Chang, D. Huh, X. Zhu, B. Cho and G. D. Smith, Microfluidic gravity pump with constant flow rate, US Pat. 7704728, The University of Michigan, 2010.
- 50 S. J. Kim, X. Zhu and S. Takayama, Gravity-Driven Fluid Pumping and Cell Manipulation, in *Microtechnology for Cell Manipulation and Sorting*, Springer International Publishing, 2017, pp. 175–192.
- 51 D. W. Lee, S. H. Yi, B. Ku and J. Kim, A pumpless perfusion cell culture cap with two parallel channel layers keeping the flow rate constant, *Biotechnol. Prog.*, 2012, 28(6), 1466–1471.
- 52 D. C. Stillinger, M. J. Head, K. N. Helland and C. W. Van Atta, A closed-loop gravity-driven water channel for density-stratified shear flows, *J. Fluid Mech.*, 1983, 131, 73–89.
- 53 S. K. Ali and J. F. Foss, A computer controlled, finite duration “steady” flow facility, *Exp. Fluids*, 1995, 19(4), 250–254.
- 54 D. G. Belair, et al., Human vascular tissue models formed from human induced pluripotent stem cell derived endothelial cells, *Stem Cell Rev.*, 2015, 11, 511–525.

