

MIT Open Access Articles

Digital optical phase conjugation for delivering two-dimensional images through turbid media

The MIT Faculty has made this article openly available. **Please share** how this access benefits you. Your story matters.

Citation: Hillman, Timothy R., Toyohiko Yamauchi, Wonshik Choi, Ramachandra R. Dasari, Michael S. Feld, YongKeun Park, and Zahid Yaqoob. "Digital Optical Phase Conjugation for Delivering Two-Dimensional Images through Turbid Media." *Sci. Rep.* 3 (May 29, 2013).

As Published: <http://dx.doi.org/10.1038/srep01909>

Publisher: Nature Publishing Group

Persistent URL: <http://hdl.handle.net/1721.1/88218>

Version: Final published version: final published article, as it appeared in a journal, conference proceedings, or other formally published context

Terms of use: Creative Commons Attribution-NonCommercial-NoDerivs 3.0





Digital optical phase conjugation for delivering two-dimensional images through turbid media

Timothy R. Hillman¹, Toyohiko Yamauchi², Wonshik Choi³, Ramachandra R. Dasari¹, Michael S. Feld^{1*}, YongKeun Park⁴ & Zahid Yaqoob¹

¹G. R. Harrison Spectroscopy Laboratory, Massachusetts Institute of Technology, 77 Massachusetts Ave., Cambridge MA 02139, ²Hamamatsu Photonics, Hamamatsu City, Japan, ³Department of Physics, Korea University, Seoul 136-701, Republic of Korea, ⁴Department of Physics, Korea Advanced Institute of Science and Technology, Daejeon 305-701, Republic of Korea.

Received
20 February 2013

Accepted
10 May 2013

Published
29 May 2013

SUBJECT AREAS:
BIOPHOTONICS
IMAGING AND SENSING
OPTICAL IMAGING
BIOMEDICAL ENGINEERING

Correspondence and requests for materials should be addressed to Y.K.P. (yk.park@kaist.ac.kr) or Z.Y. (zyaqoob@mit.edu)

* Deceased.

Optical transmission through complex media such as biological tissue is fundamentally limited by multiple light scattering. Precise control of the optical wavefield potentially holds the key to advancing a broad range of light-based techniques and applications for imaging or optical delivery. We present a simple and robust digital optical phase conjugation (DOPC) implementation for suppressing multiple light scattering. Utilizing wavefront shaping via a spatial light modulator (SLM), we demonstrate its turbidity-suppression capability by reconstructing the image of a complex two-dimensional wide-field target through a highly scattering medium. Employing an interferometer with a Sagnac-like ring design, we successfully overcome the challenging alignment and wavefront-matching constraints in DOPC, reflecting the requirement that the forward- and reverse-propagation paths through the turbid medium be identical. By measuring the output response to digital distortion of the SLM write pattern, we validate the sub-wavelength sensitivity of the system.

Optical imaging and spectroscopy of biological cells and tissue provide a tremendous amount of information that can be utilized for diagnostic and therapeutic purposes. However, the sample depth from which the information can be successfully retrieved is fundamentally limited by multiple scattering effects, mainly due to refractive index variations on a length scale comparable to the optical wavelength. Scattering is the dominant extinction process, accounting almost exclusively for the limited imaging depth range. For example, the absorption coefficient μ_a of gray matter is only 0.2 cm^{-1} , yet the scattering coefficient μ_s for the same tissue is equal to 77 cm^{-1} at the wavelength of 800 nm ¹.

One class of techniques to suppress sample turbidity involves wavefront shaping – the manipulation of the optical wavefront prior to sample illumination, pre-scrambling it in such a way to counter the effects of scattering². The value of such methods is attested by recent work demonstrating a high degree of control over the scattered optical wavefields in space³, time^{4–6}, wavelength^{7,8}, polarization^{9,10}, transmitted energy¹¹, and in the near-field (at sub-wavelength scales)¹². Recently, it has also been shown that the penetration depth of optical coherence tomography can be increased using wavefront shaping¹³. Other approaches utilize the transmission matrix, the linear transformation relating the incident wavefields to transmitted wavefields, in order to deliver optical images through highly turbid media^{14–16}.

Optical phase conjugation (OPC) is an elegant technique for reversing the effects of multiple scattering. Its first demonstration of image transmission through diffusing media was performed using photorefractive crystals as the phase-conjugate mirrors¹⁷. Our group collaborated in using conventional, crystal-based OPC to achieve turbidity suppression in biological samples¹⁸. A three-dimensional hologram of the light scattered from the sample was recorded in the crystal, which permitted a “time-reversed” version of the original scattered wave to be generated when the crystal was illuminated with an appropriate “conjugate-reference” beam. The time-reversal property was most dramatically evidenced when the phase-conjugate reconstructed wave encountered the turbid sample, leading to the reconstruction of the original, complex sample structure.

Recent advances in OPC utilize a digital implementation of the technique for turbidity suppression in biological tissues^{19,20}. In digital optical phase conjugation (DOPC), the “phase-conjugate mirror” operation is performed by two distinct devices, a *sensor* and an *actuator*¹⁹. The former is used to acquire the complex-scalar




Figure 2 | Detected intensity reconstruction in output (CCD2) plane. In (a), the target was included in the beam path. It was illuminated (in reverse) by five different speckle patterns (generated by random SLM patterns), and the average result is shown. In (b)–(d), the target was removed from the beam path (flipped down), in order to generate a reconstruction by DOPC alone. Considering the individual cases: (b) No SLM control, resulting in turbid sample illumination with a plane wave; (c) Optimized SLM control in the absence of “fine-tuning”; (d) Optimized, fine-tuned SLM control. Scale: each element-5 bar has (long-dimension) length 1.57 mm. Consistent but arbitrary intensity units are utilized in the images.

but the latter by using wavefront control to reconstruct the image utilizing the full power of the beam (and no target mask).

We address the sensitivity of the system to systematic distortions of the phase-conjugate SLM write patterns. As noted previously, precise control of the pattern is a fundamental advantage of the digital form of OPC. Specifically, we observe the effect of translating the reconstruction patterns by one, two, and three pixels from the optimized location (Fig. 3). A one-pixel shift corresponds to 20- μ m translation in the SLM plane, but due to demagnification of the beam, only approximately 100-nm lateral beam translation upon encountering the sample. Yet even a two- or three-pixel translation, corresponding to less than half of one wavelength at the sample, is sufficient to destroy the quality of the reconstruction almost entirely, demonstrating the required precision of the phase-conjugate reconstruction for highly turbid samples. This observation can be supported by inspecting a plot of the modulus of the complex autocorrelation function of the optimized phase pattern written to the SLM (see Supplementary Fig. 5).




Figure 3 | The “optimum” reconstruction is the same as in Fig. 2(d). The other reconstructions are generated by translating the SLM pattern by the specified number of pixels in the same direction. The effective translation (mismatch), Δx , of the reconstruction beam with respect to the scattering sample is indicated, subject to the qualification given in the text. The value of η corresponding to each reconstruction is calculated, also subject to a qualification given in the text. The image SNR is also calculated by comparing the reconstructed image with the image obtained in the absence of the scattering medium. The images are plotted using a linear grayscale, with intensity value 0 as the lower limit. Scale: each element-5 bar has (long-dimension) length 1.57 mm.

Discussion

To systematically address the fundamental factors that limit the quality of the reconstruction, we consider the wavefield scattered by the sample at the acquisition stage. It can be expressed as a two-dimensional field vector, $\mathbf{S} = (S_1, S_2)$, corresponding to two orthogonal polarization states. However, only a limited portion of this scattered wave, for one polarization state, is collected by the optical system, and therefore is available to generate the reconstruction. The holographically recorded field can be denoted as

$$\mathbf{S}_R = \begin{cases} (S_1, 0), & \text{if } \Omega \in \Omega_R, \\ \mathbf{0}, & \text{otherwise,} \end{cases}$$

where Ω_R is the accessible solid-angle range. The quality of the reconstruction will depend on the relative optical powers of the total “signal” and “noise” contributions of \mathbf{S}_R , defined with respect to the “target” signal \mathbf{S} . We define the *reconstruction efficiency* η to be equal to the ratio of the “signal” to the “total power”, so that:

$$\eta = \frac{|\langle \mathbf{S}, \mathbf{S}_R \rangle|^2}{PP_R}.$$

The inner product used in the equation is defined:

$$\langle \mathbf{S}, \mathbf{T} \rangle = \iint [S_1 T_1^* + S_2 T_2^*] d\Omega,$$

where the differential solid angle $d\Omega = \sin\theta d\theta d\phi$ in a spherical coordinate system, and the optical powers of the scattered and reconstruction waves are equal to $P = \langle \mathbf{S}, \mathbf{S} \rangle$ and $P_R = \langle \mathbf{S}_R, \mathbf{S}_R \rangle$, respectively. In our case, for which the components of \mathbf{S}_R are either equal to those of \mathbf{S} or zero, the reconstruction efficiency can be expressed as:

$$\eta = \frac{P_R}{P}.$$

Even though η is defined at the acquisition stage, it legitimately describes (or sets an upper limit to) the quality of the final reconstruction, since the DOPC signal is generated based on the acquired hologram.

We assume that the scattering medium thoroughly scrambles the incoming signal, so that the fraction of the scattered wave’s total power that is incident upon CCD1 is essentially independent of the specific structural form of the wave that is incident upon the medium. Therefore, the complex correlation coefficient between \mathbf{S} and \mathbf{S}_R , given by $\gamma = \langle \mathbf{S}, \mathbf{S}_R \rangle / \sqrt{PP_R}$, and hence also $\eta = |\gamma|^2$, will remain constant over the ensemble of possible incoming-signal distributions. We also assume that the spatial distribution of the “noise” component of the signal at the *reconstruction* stage is essentially independent of its particular input distribution. This means that



the background noise level captured by CCD2 at this stage is dependent only on the *total* power of the reconstruction wave incident on the reverse side of the scattering medium. Taking these statements together, we conclude that the ratio of the “signal” power to “total” power in the reconstructed image remains constant for any original target illumination pattern. Indeed, it depends only on the fundamental characteristics of the recording system such as the numerical aperture for collection of the scattered light (and to some degree, the angular scattering characteristics of the sample). The ratio could also be improved by incorporating polarization-sensitivity into the system.

As we have noted, the value of η calculated according to fraction of the scattered light collected at the acquisition stage is only an upper limit to the actual “reconstruction efficiency” (at CCD2), since the “true” quantity depends on our ability to effectively construct the phase-conjugate beam and back-propagate it through the system. We observe that the aim of the correction steps, detailed in the Supplementary information, was to precisely match the phase profiles of scattered and reconstruction wavefronts up to phase conjugation. A technical factor that diminishes the effective value of γ is the fact that we utilize phase control only (not amplitude control) in generating the reconstruction wave. Assuming that the amplitude distribution across the recording-plane field of view obeys Rayleigh statistics, then this effect decreases γ by a factor of $2/\sqrt{\pi} = 1.13$, with a consequent reduction in η by the square of this value.

Having established that η is not dependent on the structure or power of the original target beam, but only upon the system characteristics, we now consider the *visibility* of a target reconstruction. The visibility can be defined as the extent to which the reconstruction (at CCD2) is distinguishable from the background noise, the ratio of the average image *intensity* level (above the noise floor) in the region corresponding to the transparent area of the mask, to the background intensity level. It depends primarily on the spatial *extent* of the reconstruction pattern (or the mask area). This is because a structure-independent signal power level (relative to the background) is dispersed over this area. Accordingly, the broader the target pattern, the greater the system η required to reconstruct it to high fidelity. This observation explains why a single point target can be much more easily reconstructed than a complex image: even if the reconstruction efficiency η is low, the visibility may still be extremely high given that the signal component of the reconstruction is concentrated to a focused spot, but the noise component is spread over a large scattering solid angle. The high-quality reconstruction obtained in Ref. 18 can be accounted for due to the high system η : the scattering medium was in direct contact with the recording crystal, so that virtually all of the forward-scattered light contributed to the hologram. Moreover, the holographic reconstruction provided by the 3D recording medium was a faithful representation of the scattered wavefield. These system characteristics permitted a high-visibility reconstruction of the large target area to be obtained.

The unusual feature of our DOPC system, that image fidelity depends on the spatial extent of the target pattern, is similar in nature to that of the related technique of Ref. 3. In this paper, when an SLM-pattern optimization routine is used to generate *multiple* high-intensity focused spots in a recording plane located behind a scattering sample, then the *total* optical power of the signal-carrying component of the light is the same as for the generation of a single spot, but it is divided between the spots.

In Fig. 3, we have stated representative (effective, reconstruction-plane) values of η calculated by dividing the measured power in the “signal” region of the image (corresponding to the transparent region of the mask) to the sum of this and the measured power in the “background” region (elsewhere, including a larger field of view not included in the cropped depictions of the figures). Because the optical imaging system at CCD2 has a limited collection aperture, the contribution of light scattered at large off-axis angles is not included

in the calculation, so the reported values are overestimates. However, the error in measured background power will be a systematic (constant) scaling factor. The visibility is also reported in the figure. Because it is based on intensity measurements, not total power measurements, it can be determined more accurately. Note that since a single mask pattern is used, there is a direct correspondence between visibility and η . (It is not strict proportionality, because the term on the denominator of η is total power, not “background” or noise power.) The reconstruction quality (as evaluated by inspection, or by measured visibility), can be very good despite the fact that the η values are low (substantially less than 1 even in the optimum case). We reiterate that this is because the “signal-bearing” component of the scattered light is concentrated in a small area, whereas the “noise” component is diffused over a large area.

In summary, we have experimentally demonstrated wide-field imaging through a highly scattering sample using DOPC employing a ring interferometer configuration. The challenging optical alignment problem of matching the forward- and reverse-propagating wavefields has effectively been solved. The SLM used as the phase-conjugate-mirror “actuator” was capable of being directly imaged onto the CCD used as “sensor” in the ring interferometer, the key to the sub-wavelength accuracy attained in the optical alignment. Our work provides a means to implement a robust and practical DOPC system for biophotonic applications where multiple light scattering is a limiting factor, including optical deep-tissue imaging and photodynamic therapy.

Methods

Experimental setup. In our DOPC system, an EM-CCD camera (CCD1 in Fig. 1, Photon Max 512, Princeton Instruments, New Jersey, USA) and a SLM (Hamamatsu, LCOS-SLM X10468-02, Japan) constitute the sensor and actuator, respectively. A phase profile (pixelated over an 800×600 grid, total dimensions 16×12 mm) is applied to the beam incident on the SLM. Full phase control is permitted for each pixel independently, and performed using a computer. The liquid crystal pixels of the SLM are aligned horizontally, mandating that horizontally polarized light be incident upon it to achieve phase-only control. Highly coherent light from a tunable laser source (Xperay Tunable Laser, General Microtechnology and Photonics, Switzerland), operating at 831 nm, is coupled into a single-mode fiber that is further connected to an optical-fiber-based polarization controller. This permits the source to be easily interchanged with a visible-light fiber-coupled laser, for the purpose of system alignment. After exiting the fiber, the beam is expanded and collimated; in Fig. 1, it is shown entering the system from the top-right.

Sample preparation. The scattering layer was prepared by spray-painting ZnO particles (an average diameter of 200 nm) on a standard glass microscope cover slip. The transport mean free path of the sample was measured as 0.7 ± 0.2 μm at a wavelength of 632.8 nm and the thickness of the sample was 11.3 μm . The turbid sample was positioned at the common focal plane of the two microscope objectives.

1. Vo-Dinh, T. Biomedical photonics handbook. *Journal of Biomedical Optics* **9**, 1110 (2004).
2. Mosk, A. P., Lagendijk, A., Leroose, G. & Fink, M. Controlling waves in space and time for imaging and focusing in complex media. *Nat. Photonics* **6**, 283–292 (2012).
3. Vellekoop, I. & Mosk, A. Focusing coherent light through opaque strongly scattering media. *Optics Letters* **32**, 2309–2311 (2007).
4. McCabe, D. J. *et al.* Spatio-temporal focusing of an ultrafast pulse through a multiply scattering medium. *Nature Communications* **2**, 447 (2011).
5. Katz, O., Small, E., Bromberg, Y. & Silberberg, Y. Focusing and compression of ultrashort pulses through scattering media. *Nature photonics* **5**, 372–377 (2011).
6. Aulbach, J., Gjonaj, B., Johnson, P. M., Mosk, A. P. & Lagendijk, A. Control of light transmission through opaque scattering media in space and time. *Phys. Rev. Lett.* **106**, 103901 (2011).
7. Park, J. H., Park, C. H., Yu, H., Cho, Y. H. & Park, Y. K. Active spectral filtering through turbid media. *Opt. Lett.* **37**, 3261–3263 (2012).
8. Small, E., Katz, O., Guan, Y. & Silberberg, Y. Spectral control of broadband light through random media by wavefront shaping. *Optics Letters* **37**, 3429–3431 (2012).
9. Park, J.-H., Park, C., Yu, H., Cho, Y.-H. & Park, Y. Dynamic active wave plate using random nanoparticles. *Opt. Express* **20**, 17010–17016 (2012).
10. Guan, Y., Katz, O., Small, E., Zhou, J. & Silberberg, Y. Polarization control of multiply-scattered light through random media by wavefront shaping. *Optics Letters* **37**, 4463–4465 (2012).



11. Kim, M. *et al.* Maximal energy transport through disordered media with the implementation of transmission eigenchannels. *Nature photonics* **6**, 581–585 (2012).
12. Park, J. H. *et al.* Subwavelength light focusing using random nanoparticles. *Nature photonics* (2013).
13. Jang, J. *et al.* Complex wavefront shaping for optimal depth-selective focusing in optical coherence tomography. *Optics Express* **21**, 2890–2902 (2013).
14. Popoff, S., Lerosey, G., Fink, M., Boccaro, A. C. & Gigan, S. Image transmission through an opaque material. *Nature Communications* **1**, 81 (2010).
15. Popoff, S., Lerosey, G., Fink, M., Boccaro, A. & Gigan, S. Controlling light through optical disordered media: transmission matrix approach. *New Journal of Physics* **13**, 123021 (2011).
16. Choi, Y. *et al.* Overcoming the diffraction limit using multiple light scattering in a highly disordered medium. *Physical Review Letters* **107**, 23902 (2011).
17. Leith, E. N. & Upatnieks, J. Holographic imagery through diffusing media. *J. Opt. Soc. Am. A* **56**, 523–523 (1966).
18. Yaqoob, Z., Psaltis, D., Feld, M. & Yang, C. Optical phase conjugation for turbidity suppression in biological samples. *Nature photonics* **2**, 110 (2008).
19. Cui, M. & Yang, C. Implementation of a digital optical phase conjugation system and its application to study the robustness of turbidity suppression by phase conjugation. *Optics Express* **18**, 3444–3455 (2010).
20. Hsieh, C. L., Pu, Y., Grange, R., Laporte, G. & Psaltis, D. Imaging through turbid layers by scanning the phase conjugated second harmonic radiation from a nanoparticle. *Optics Express* **18**, 20723–20731 (2010).
21. Garibyan, O. *et al.* Optical phase conjugation by microwatt power of reference waves via liquid crystal light valve. *Opt Commun* **38**, 67–70 (1981).
22. Grange, R., Lanvin, T., Hsieh, C. L., Pu, Y. & Psaltis, D. Imaging with second-harmonic radiation probes in living tissue. *Biomedical optics express* **2**, 2532–2539 (2011).
23. Yang, Y. M., Judkewitz, B., A. D. C. & Yang, C. Deep-tissue focal fluorescence imaging with digitally time-reversed ultrasound-encoded light. *Nature Communications* **3**, 8 (2012).
24. Si, K., Fiolka, R. & Cui, M. Fluorescence imaging beyond the ballistic regime by ultrasound-pulse-guided digital phase conjugation. *Nature photonics* (2012).
25. Vellekoop, I. M., Cui, M. & Yang, C. Digital optical phase conjugation of fluorescence in turbid tissue. *Appl Phys Lett* **101**, 081108–081108–081104 (2012).
26. Vellekoop, I. M. & Aegerter, C. M. Scattered light fluorescence microscopy: imaging through turbid layers. *Opt. Lett* **35**, 1245–1247 (2010).

Acknowledgments

This work was supported by The National Institute of Biomedical Imaging and Bioengineering (9P41-EB015871-26A1), Hamamatsu Photonics (Japan), Samsung Advanced Institute of Technology, KAIST Institute for Optical Science and Technology, and the Korean Ministry of Education, Science and Technology (MEST) grant No. 2009-0087691 (BRL) and National Research Foundation (2012R1A1A1009082, 2012K1A3A1A09055128, M3C1A1-048860, 2013M3C1A3000499). TRH acknowledges support from the Singapore-MIT alliance fund (SMA2). YKP acknowledges support from TJ ChungAm Foundation. We thank Dr. Allard Mosk (University of Twente, The Netherlands) for providing the scattering samples.

Author contributions

T.R.H., Y.P. and Z.Y. designed experiments, performed research, analyzed the data and wrote the paper; T.Y. and W.C. provided advice and assistance at the early stages of the project; R.R.D. and M.S.F. provided advice and were responsible for obtaining financial support for the project. All surviving authors reviewed the manuscript.

Additional information

Supplementary information accompanies this paper at <http://www.nature.com/scientificreports>

Competing financial interests: The authors declare no competing financial interests.

License: This work is licensed under a Creative Commons Attribution-NonCommercial-NoDerivs 3.0 Unported License. To view a copy of this license, visit <http://creativecommons.org/licenses/by-nc-nd/3.0/>

How to cite this article: Hillman, T.R. *et al.* Digital optical phase conjugation for delivering two-dimensional images through turbid media. *Sci. Rep.* **3**, 1909; DOI:10.1038/srep01909 (2013).