



# MIT Open Access Articles

## *Validating Transcripts with Probes and Imaging Technology*

The MIT Faculty has made this article openly available. **Please share** how this access benefits you. Your story matters.

<b>Citation</b>	Itzkovitz, Shalev, and Alexander van Oudenaarden. "Validating transcripts with probes and imaging technology." <i>Nature Methods</i> 8.4s (2011): S12-S19.
<b>As Published</b>	<a href="http://dx.doi.org/10.1038/nmeth.1573">http://dx.doi.org/10.1038/nmeth.1573</a>
<b>Publisher</b>	Nature Publishing Group
<b>Version</b>	Author's final manuscript
<b>Accessed</b>	Thu Jun 21 02:52:20 EDT 2018
<b>Citable Link</b>	<a href="http://hdl.handle.net/1721.1/71013">http://hdl.handle.net/1721.1/71013</a>
<b>Terms of Use</b>	Creative Commons Attribution-Noncommercial-Share Alike 3.0
<b>Detailed Terms</b>	<a href="http://creativecommons.org/licenses/by-nc-sa/3.0/">http://creativecommons.org/licenses/by-nc-sa/3.0/</a>

# Validating Transcripts with Probes and Imaging Technology

Shalev Itzkovitz<sup>1,2</sup>, Alexander van Oudenaarden<sup>1,2,3†</sup>

<sup>1</sup>Department of Physics, Massachusetts Institute of Technology, Cambridge MA 02139, USA.

<sup>2</sup>Department of Biology, Massachusetts Institute of Technology, Cambridge MA 02139, USA.

<sup>3</sup>Koch Institute for Integrative Cancer Research, Massachusetts Institute of Technology, Cambridge, MA 02139, USA.

†To whom correspondence should be addressed: E-mail: [avano@mit.edu](mailto:avano@mit.edu)

## **Abstract**

High throughput gene expression screens provide a quantitative picture of the average expression signature of biological samples. However, the analysis of spatial gene expression patterns with single cell resolution requires quantitative *in-situ* measurement techniques. Here we describe recent technological advances in RNA fluorescent *in-situ* hybridization (FISH) techniques that facilitate detection of individual fluorescently labeled mRNA molecules of practically any endogenous gene. These methods, which are based on advances in probe design, imaging technology, and image processing, enable the absolute measurement of transcript abundance in individual cells with single-molecule resolution.

High-throughput measurements of gene expression on a genomic scale using microarray technology or high throughput sequencing contributed tremendously to our understanding of how genetic networks coordinately function in normal cells and tissues and how they malfunction in disease. Such measurements allow one to infer the function of genes based on their co-expression patterns<sup>1</sup>, to detect which genes have altered expression in disease<sup>2</sup>, and to identify expression signatures that are predictive of cancer progression<sup>3,4</sup>. However, the variability in single cell gene expression in most biological systems and especially in tissues and tumors suggests that bulk transcriptome measurements should be complemented by techniques aimed at characterizing gene expression programs in individual cells<sup>5</sup>. This review will describe advances in single-molecule transcript imaging, which yield integer counts of transcripts in single cells in suspension and within intact tissues.

## Biological samples are inherently heterogeneous

Bulk transcriptome measurements inform on the average gene expression in a sample. Thus in a heterogeneous sample, containing several cell types with different gene expression signatures, only the most abundant signature will be captured. Such heterogeneity is present in practically any biological sample. Even bacterial and yeast cells that are derived from isogenic monoclonal populations have been shown to exhibit pronounced cell to cell variation in the expression of many genes, stemming from stochastic events such as bursting transcriptional dynamics and cell cycle dependence<sup>6</sup>. Expression heterogeneity is even more pronounced in tissues, which are usually composed of several types of cells with profoundly different gene expression programs. In many epithelial tissues, such as the skin and the intestine, there is a clear hierarchical partition into stem cells and diverse differentiated epithelial progenies, each of which displays distinct phenotypic and morphological features. The precise location of cells within the tissue translates to constant changes in the levels of niche-secreted morphogens, which give rise to position-modulated gene expression programs. Thus two adjacent cells could harbor dramatically different expression programs.

Solid tumors represent a particular case of cell heterogeneity. Most solid tumors consist of a mixture of cancer and stromal cells. Additionally, cancer cells often show profound diversity not only in their transcript content but also in their genotype. This diversity stems from increased mutation rates, rapid cell proliferation and spatially varying selection forces. Single cells from a wide range of colorectal cancer cell lines change their chromosomal copy number on average once every five cell divisions *in vitro*<sup>7</sup>, and there is a dramatic heterogeneity in copy numbers observed in tissue cross-sections<sup>8,9</sup>. A bulk measurement of gene expression in tumors only captures the most dominant tumor clone and is masked by stromal signals, therefore providing only partial information on the different expression signatures co-existing in the tumor. Moreover, some tumors have been shown to contain a minority of cells with unlimited proliferative capacity and increased oncogenic ability termed cancer stem cells<sup>10</sup>. Bulk measurements of tumors cannot capture the expression signature of such minor populations.

Attempts to address this underlying heterogeneity are based on enriching for certain subpopulations from tissues or tumors. An example of such enrichment is the use of laser capture microdissection, enabling extracting regions within a tissue based on morphological features or gene expression markers<sup>11</sup>. This technique is frequently applied to selectively collect tumor cells rather than a mixture of tumor and stroma cells for down-stream expression analysis. Alternatively a fluorescence activated cell sorter can be used to specifically isolate cells expressing a small number of defined gene expression markers. These cells can then be used for either bulk measurements or for single cell measurements that require cell lysis, such as quantitative RT-PCR<sup>12</sup> or digital RT-PCR<sup>13</sup>. Similarly, sorting GFP positive cells from transgenic mice expressing GFP under the control of a cell-type specific promoter of interest enables enriching for the cells expressing this gene and allows characterization of the expression program of these particular cells, as has been done with tissue stem cells<sup>14</sup>. The main drawback of these methods is the fact that they involve dissociating the tissue and thus result in the complete loss of spatial information. Moreover, specific expression of a single gene in only one cell type is uncommon, and thus the enriched cell population would generally still consist of a mixture of

different cell populations. In addition, the variability resulting from the reverse transcription and the exponential PCR amplification steps often limits the sensitivity and reproducibility of these approaches<sup>12</sup>. Thus quantitatively unraveling the gene expression programs of single cells within heterogeneous samples while preserving spatial context requires transcript measurement methods in intact tissues.

### **Traditional approaches for measuring *in-situ* gene expression**

The traditional method for measuring gene expression in individual cells in intact tissues is RNA *in-situ* hybridization (ISH) and its variant RNA fluorescent in-situ hybridization (FISH)<sup>15</sup>. FISH methods have originally been developed for DNA analysis<sup>16</sup>, and are based on the specific binding of long fluorescently labeled oligonucleotide probes to their complementary sequences in fixed and permeabilized samples. While there are many nuances to FISH techniques most entail the following steps (Fig. 1). Tissues or cells are fixed, permeabilized, and then hybridized with fluorescently labeled oligonucleotide probes. Probes are either directly coupled to fluorescent molecules, or alternatively coupled to haptens, typically biotin or digoxigenin (DIG), to which either avidin or anti-DIG antibody bind respectively. In turn these are either directly conjugated to fluorophores, or to enzymes such as alkaline phosphatase or horseradish peroxidase, which generate either fluorescent (in FISH) or chromogenic (in ISH) products from specific added substrates. Indirect labeling offers the advantage that signal amplification can be achieved via the use of primary and secondary antibodies, of which only the secondary antibody elicits a light emitting reaction<sup>17</sup>.

Coupling fluorophores to the oligonucleotide probes has traditionally been achieved by enzymatic means such as nick translation or *in-vitro* transcription, methods in which dye-modified nucleotides are stochastically incorporated along the probe (**Fig. 2a**). While detection of single molecules using traditional FISH has been reported for a specific probe in *Drosophila*<sup>18</sup>, in most cases traditional ISH and FISH techniques can only give qualitative information on gene expression. One drawback is the random distribution of dyes along the linear sequence, which has been shown to impede the hybridization quality by destabilizing the probe-target duplex, often leading to mutual quenching of adjacent dyes<sup>19</sup>. Also, long probes show poor permeability, high background and low sensitivity. Since it has been estimated that more than 80% of yeast genes are expressed at lower than two mRNA copies per cell<sup>20</sup>, it would be impossible to study the majority of the eukaryotic transcriptome with traditional FISH.

### **Single-molecule transcript imaging**

Recent improvements in probe designs, imaging technology and image processing software have emerged that enable highly specific and robust hybridization of fluorescently labeled probes to an arbitrary transcript of interest with high spatial localization, yielding diffraction limited spots under a

fluorescence microscope (Table 1). Thus an integer number of transcripts in individual cells can be determined by simply counting fluorescent dots. Moreover, the precise sub-cellular localization of the individual transcripts can be detected.

#### *Short probes labeled with multiple fluorophores*

Robert Singer and colleagues pioneered single mRNA molecule imaging techniques<sup>21</sup>. The key improvement was the replacement of long probes with 50 bp probes, that are complementary to sequential parts of the target mRNA and are each coupled to typically 3-5 fluorescent dyes at predefined positions (**Fig. 2b**). A minimal spacing between incorporated dyes along the linear oligonucleotide sequence was set to prevent quenching of adjacent dyes<sup>22</sup>. Coupled probes were separated from non-coupled probes and free dyes using chromatography. Singer and colleagues optimized the GC content of all probes to be as close as possible to 50%, and imaged the hybridized samples with a fluorescence microscope that captured stacks of images every fraction of microns, a high numerical aperture objective and a low-noise CCD camera. These improvements yielded three dimensional digital images in which the specific accumulation of fluorescent molecules in the small volume occupied by a single mRNA molecule appeared as diffraction limited spots.

Several research groups have since applied the use of multiply labeled probes for single mRNA counting to provide a detailed characterization of transcripts distribution in yeast<sup>23-25</sup> and in mammalian cells<sup>26</sup>. In mammalian tissue, this method did not yet achieve single transcript resolution but enabled the detection of transcription sites in paraffin embedded human tumors<sup>27</sup>. But the approach also has some drawbacks, mainly a high variability in the number of probes bound to the target<sup>21</sup>. While ideally each spot would be composed of the same number of probes, in fact most fluorescent spots were estimated to originate from only one or two probes. This made it difficult to differentiate the true specific binding to the legitimate target from non-specific binding. Since an inefficient binding of one or two probes can lead to a high variability in spot intensity when only a handful of probes are used, it is important to carefully select the probes using this method. An additional source of variability using probes designed to be coupled to several fluorescent molecules is in the number of fluorescent molecules actually coupled to each probe - it is difficult to separate fully coupled probes from partially coupled ones.

#### *Short probes with single fluorophores*

Raj et al.<sup>28</sup> modified the Singer method to create probe libraries consisting of many (typically between 48 and 96) short, 17-22 bp oligos labeled with only a single fluorophore at their 3' termini<sup>28</sup> (**Fig. 2c**). This allowed more efficient purification and the variability in spot intensity caused by inefficient binding of any single probe was lower when multiple probes were used. By optimizing probe design to have a uniform GC content of around 45% and a minimum gap of three nucleotides between successive probes, robust specific hybridization was achieved with a set of 48 probes or less. Raj et al.<sup>28</sup> demonstrated

simultaneous transcript counting of three transcripts coupled to distinct fluorophores. The use of a specialized mounting medium containing an oxygen scavenging system inhibited oxygen-dependent, light initiated pathways that destroy fluorophores, thus increasing the photostability of the dyes. This is particularly important in the optical setup in which the same field of view is illuminated multiple times as the optical sections are gathered<sup>22</sup>.

Several research groups have successfully applied singly labeled probes to a wide range of samples, ranging from yeast<sup>29</sup> and mammalian cells<sup>28,30</sup>, to *Drosophila*<sup>28</sup> and *C. elegans* embryos<sup>31</sup>. The approach is appealing mainly for its simplicity and generality, aided by a simple web interface to design optimal probes for arbitrary sequences ([www.singlemoleculefish.com](http://www.singlemoleculefish.com)). Probe libraries are typically prepared with a 96-position DNA synthesizer, pooled and then simultaneously coupled to a fluorophore of choice. This format facilitates additional flexibility by enabling coupling of the same library to several different fluorescent dyes for simultaneous hybridization to other genes of interest.

A current limitation is the difficulty in detecting short transcripts. Xie and coworkers recently demonstrated single mRNA detection with only one 20bp fluorescently labeled probe in *E. coli*<sup>32</sup>. The ability to detect single transcripts in mammalian cells using one probe is much more challenging due to the larger volumes and increased off-target sequences involved. *In-situ* detection of short transcripts using oligonucleotide probes can be achieved by using modified nucleic acids to increase specificity and by applying signal amplification to increase sensitivity. Next we will discuss these two approaches.

#### *Short probes with modified backbones*

When the target sequences are too short, one runs into specificity problems, as any single probe has a non-negligible probability to bind a different target. Increasing the sensitivity and specificity of individual probes could enable the detection of very short transcripts or of single microRNA molecules, which typically span 19-24 bps<sup>33</sup>, as well as allow distinguishing between transcript variants that differ by single nucleotides<sup>34</sup>. This can be achieved with probes containing modified nucleic acids. The most commonly used modifications are peptide nucleic acids (PNA) and locked nucleic acids (LNA). PNA have an uncharged peptide-like backbone<sup>35</sup> and therefore hybridizes more stably to RNA compared to DNA probes. LNA is a 2'-O, 4'-C-methylene-linked ribonucleotide derivative of RNA, enabling more specific hybridization with RNA and DNA, compared to DNA or RNA probes. PNA probes have been used to detect telomeres - repetitive hexamer sequences at chromosome ends<sup>36</sup>, whereas LNA probes facilitated the detection of microRNAs<sup>33,37</sup>, but with the use of signal amplification.

#### *Signal amplification of single-molecule probes*

Detection of short transcripts is still hampered by insufficient fluorescence from a single bound probe. Amplifying the fluorescent signal can solve this sensitivity limitation. Single microRNA molecules have

been detected *in situ* using a single LNA probe and an alkaline phosphatase–based signal amplification<sup>33</sup>. Larsson *et al.*<sup>34</sup> introduced the use of padlock probes that can distinguish transcripts that differ by only a single base pair. The authors first reverse-transcribed the mRNA into cDNA using LNA primers, then hybridized linear padlock probes to juxtaposed segments of the target sequence, enzymatically ligated them and used them as templates for rolling circle amplification by Phi29 DNA polymerase (**Fig. 2d**)<sup>34</sup>. This created a single strand of DNA containing tandem repeats of the padlock probe sequence that, after hybridization to fluorescently labeled probes, yielded bright diffraction-limited spots. Larsson *et al.*<sup>34</sup> demonstrated multiplex detection of up to three genes in cells and frozen mouse embryonic tissue.

In the branched DNA approach<sup>38</sup> hybridization is performed with four distinct probe sets: a gene-specific probe set composed of ten or more oligonucleotide probe pairs that are complementary to the target sequence, a pre-amplifier probe that hybridizes to gene-specific probes, multiple amplifier probes that hybridize to the pre-amplifier probe and labeled probes that attach to the amplifier probes. The resulting construct yields bright concentrated fluorescence (**Fig. 2e**). This technology is commercially named QuantiGene ViewRNA<sup>38</sup>.

### *Spectral barcoding*

The number of spectrally resolvable fluorophores, typically three, limits the number of simultaneously measured genes in transcript imaging techniques. ‘Spectral barcoding’ is an approach developed by Singer and colleagues to increase the number of simultaneously detected transcripts<sup>21,39</sup>. The technology, which was based on a similar method used for DNA FISH<sup>40,41</sup>, entailed the division of the probe set that is complementary to a given transcript of interest into groups, each of which is coupled to a different fluorophore. By precisely registering the images from the different fluorescence channels one can determine not only whether a spot is present or not, but also how many colors compose it. With  $n$  spectrally resolvable fluorophores one can in principle achieve  $2^n - 1$  different probe color combinations. This limit could be increased if the detection method is sensitive enough to detect not only whether a given fluorophore is present or absent at a diffraction limited spot, but to also estimate the number of oligonucleotides coupled to the specific fluorophore of interest<sup>39</sup>.

### *Single-molecule transcript imaging in live cells*

A limitation of transcript imaging using FISH approaches is that these require the samples to be fixed. The ability to temporally measure gene expression in living cells offers a rich source of additional information, including a detailed description of the life of an mRNA molecule, from transcription through intra-cellular trafficking, translation and degradation. Transcript imaging in live cells is considerably more challenging than in fixed cells, both because of the gentler chemical conditions that must be applied to preserve cell integrity, and the more intricate image processing steps, requiring tracking the trajectory of a transcript with time, a task that is often confounded by the high diffusion

rates of molecules and the proximity to other molecules<sup>42</sup>. Two main technologies enable real-time measurements of transcript levels in single cells – the MS2 method and molecular beacons.

The coat protein of the bacteriophage MS2 binds a specific RNA hairpin loop. In the MS2 system, a fluorescent reporter protein gene such as *GFP* is fused to the MS2 coat protein gene, and several successive target RNA hairpin–encoding sequences are inserted into the 3′ untranslated region of a target gene of interest. Simultaneous binding of several MS2-GFP proteins to the target transcript yields a concentrated fluorescence signal. This method has been used to follow the dynamics of gene expression in bacteria<sup>43</sup>, yeast<sup>44</sup> and mammalian cells<sup>45,46</sup>, achieving single-transcript resolution. Moreover, a general method to tag and image any mRNA in yeast has recently become available<sup>47</sup>. A limitation of the MS2-GFP approach is the high background fluorescence associated with the MS2-GFP molecules that are not bound to the target RNA. Split-GFP approaches<sup>48–51</sup> alleviate this limitation, as each fragment is attached to a different RNA-binding protein that targets a distinct RNA motif. By constructing these motifs in tandem at the 3′ untranslated region of the target, an intact fluorescent GFP is assembled only when bound to the target transcript. The main limitation of the MS2 and split-GFP methods is the requirement to generate transgenes for the *GFP* construct. Yoshio Umezawa and colleagues recently demonstrated the ability to target endogenous mRNA by fusing the GFP fragments to the RNA-binding domain of the Pumilio protein, which can be designed to facilitate specific binding to an arbitrary RNA sequence<sup>51</sup>. A remaining drawback is that the binding of multiple proteins to the target mRNA could modify its intracellular dynamics, thus providing a nonrepresentative picture of the endogenous conditions.

Another technique for following individual transcripts in live cells uses molecular beacons<sup>52,53</sup> – nucleic acid probes that form hairpins coupled to a fluorophore and a quencher on opposite ends. Specific hybridization to the target sequence causes a conformational change that physically separates the quencher and fluorophore, resulting in light emission. Transcript detection in live cells is more challenging owing to their decreased stability caused by intracellular RNases and their non-homogenous intra- cellular distribution<sup>49</sup>. Probes with 2′-O-methylribonucleotide, which is not a target of RNase H, have been shown to alleviate this problem<sup>54–56</sup>. Molecular beacons require microinjection into cells, a procedure with high yield but one that may affect cell viability and can often result in a rapid drift of the microinjected probes into the nucleus. As an alternative, Santangelo *et al.*<sup>56</sup> used reversible permeabilization of the plasma membrane with pore-forming toxins, such as streptolysin-O, to deliver multiply labeled tetravalent RNA probes with minimal cytotoxicity. These probes, which are complementary to the target RNA and labeled with three fluorescent molecules each, tetramerize through their additional binding to streptavidin, thus yielding increased accumulation of fluorophores at the target transcript and allowing visualization of RNA dynamics<sup>56</sup>.

### **Imaging technology for single-molecule FISH**

Detection of single mRNA molecules imposes specific requirements on the imaging platform in terms of spatial resolution and sensitivity. For lens-based systems the maximum resolution can be estimated by



the Rayleigh criteria, which says the resolution is equal to  $(0.61)\lambda/NA$ , in which  $\lambda$  is the illumination wavelength, and NA is the numerical aperture of the lens. In practice, this limits the lateral resolution to 200–400 nm. Optimizing spatial resolution typically requires the use of a high-numerical-aperture oil-immersion lens and immersion oil that exactly matches the refractive index of the lens and the cover glass. Achromat or apochromat objectives minimize chromatic aberrations and are thus necessary when multiplexing different fluorophores<sup>42</sup>. Resolving diffraction-limited volumes also requires using a camera with sufficient spatial resolution that does not compromise the optimal optical resolution. CCD cameras have pixel sizes of 2–40  $\mu\text{m}$ , which translates to a size of 20–400 nm at the image plane with a 100 $\times$  objective depending on the optical setup. To achieve optimal resolution, it is crucial to determine the size of the image being projected onto the camera and verify that there are 2.5–3 pixels per unit resolution. If pixel binning is used to increase the signal at the camera, the number of required pixels must be multiplied accordingly. Exposure times should be optimized: long exposure times lead to higher precision in transcript localization but could accelerate sample bleaching<sup>57</sup>.

Image signal is often masked by noise stemming from both instrument-related factors such as dark current, pixelation noise and CCD readout noise but mostly from out-of-focus light from cellular autofluorescence<sup>57,58</sup>. Maximizing signal-to-noise ratio requires increasing photon output of the fluorescent labels and limiting background noise. Wide-field epi-illuminated microscopes use a broad-band light source with a spectrum from 250 nm to 1,100 nm. Excitation filters can narrow this band to the fluorophore excitation wavelength, but autofluorescence always remains. Collecting light from a wide field also results in an increase in background. Confocal microscopes have the advantage of using a pinhole or slit to block detection of light from outside the focal region, thus potentially increasing relevant fluorescence and decreasing autofluorescence. Disadvantages of confocal microscopes over epi-fluorescence microscopes include their higher cost, the lower throughput owing to the necessity to scan the image (partly alleviated by using spinning disk confocal microscopes), more stringent safety requirements and the additional descanning optics necessary to guide the detected light to the detector, which can limit transmission efficiency. Multiplex detection of several fluorophores can be achieved by using filter wheels for epifluorescence microscopes and multimode lasers for confocal microscopes. Instrument-related noise, such as dark current can be substantially minimized by cooling the CCD chip to  $-80\text{ }^{\circ}\text{C}$ <sup>58</sup>.

## Data analysis

Extracting gene expression from single-molecule FISH experiments requires automatic detection of spots representing single mRNA molecules in digital three dimensional image stacks. The image processing steps usually include image enhancement - either deconvolution or high pass filtering, followed by thresholding and localization of connected components<sup>21,22,28,31,34</sup>. A point source in the illuminated image plane is expanded and distorted by the microscope and the detector optics resulting in a blurred three dimensional version, called the point spread function (PSF). Because a fluorescence emission pattern can be treated as a sum of point sources, a recorded image is a convolution of the real image

and the PSF. When an accurate description of the PSF is available one can use deconvolution algorithms to reconstruct the original image<sup>59</sup>. Such algorithms deblur images and reassign photons emanating from out-of-focus z-dimension planes back to their original plane. Successful deconvolution requires precise measurements of the PSF, which is typically achieved by imaging sub-diffraction fluorescent beads<sup>21,22</sup>. An alternative to deconvolution is to filter the image stack with a three dimensional high-pass filter (such as a 3D Laplacian of Gaussian filter) to enhance features of the relevant spot dimension<sup>28,31</sup>. The next step after image enhancement usually consists of applying a threshold to the enhanced image yielding a binary image, in which connected components can be localized. Since a connected component spans more than a single pixel its centroid achieves sub-pixel resolution for the localization of the source. The value of the image thresholds can affect the number of spots detected and a reasonable choice for this value is one at which the number of detected spots is least sensitive to threshold selection<sup>28,31</sup>. Alternatively least-square Gaussian fitting algorithms can be used to localize the point source in a gray scale image<sup>57,60</sup>.

## **Outlook**

One appealing application of single transcript imaging is the validation of regulatory interactions. A traditional approach entails engineering mice in which a regulatory gene of interest is deleted, and searching for putative target genes with modified expression levels<sup>61</sup>. Single-molecule FISH can enable detecting such targets in wild-type tissue by hybridizing to a sample probes for the regulatory gene and for a putative target gene. A measured positive correlation in the transcript abundance could imply either a direct regulation or alternatively regulation by a common upstream component. Single-molecule transcript imaging can also provide valuable information on the behavior of network motifs, modular circuit components such as feedback and feedforward loops which are highly abundant in transcriptional networks and often comprise only a few genes<sup>62</sup>. The simultaneous quantitative *in-situ* measurements of a handful of different transcripts can shed light on the behavior of these motifs within their tissue context. In a tumor single-molecule transcript imaging can highlight the role of transcriptional heterogeneity in tumor progression and the relation between spatial context and phenotypic states of cells, represented by their expression signatures.

Single-molecule transcript imaging techniques can be combined with high-throughput expression analysis in two complementary ways. One would be to start out with large gene expression screens that would suggest putative genes of interest, the detailed *in-situ* expression of which would then be described using single-molecule FISH. An alternative approach would be to start with an unbiased mapping of a tissue using a panel of single-molecule FISH probes to detect an interesting expression pattern in terms of spatial distribution within a tissue or an unusual co-expression pattern of a few genes in isolated cells. One could then enrich for such cells using FACS or laser capture and extend this core gene expression signature with high-throughput genome-wide expression measurements. This approach could provide a detailed description of rare cell populations residing within a tumor, such as putative cancer stem cells

A technical limitation of single-molecule transcript imaging is the inability to spatially resolve single molecules when they are closer than the diffraction limit, typically 200 nm. This could be a significant problem in highly expressed genes such as ribosomal components especially in smaller organisms such as yeast<sup>25</sup>, and when mRNA molecules are physically localized in transport particles<sup>63</sup>. Techniques that can address this limitation are sub-diffraction-limit microscopy methods such as STED<sup>64</sup>, STORM<sup>65</sup> and PALM<sup>66</sup>, which enable resolving fluorescent molecules with nanometer resolution<sup>67</sup>. While sub-diffraction microscopy outperforms other technologies in spatial resolution, enabling probing molecular structures in fixed and even live cells<sup>68,69</sup>, scaling the technology to comprehensively measure gene expression *in-situ* in many cells and in tissues is still a challenge due to long recording times, high intensity illumination, the prolonged use of which could potentially be harmful to the sample and expensive instrumentation.

Another exciting recent development is the use of probes coupled to quantum dots<sup>70-73</sup>. The brightness of quantum dots makes them especially attractive for studying tissues, where cellular autofluorescence often masks the signal. Oligonucleotide probes labeled with quantum dots have been used to detect transcripts in cells and tissues<sup>70-72</sup>, in paraffin embedded tissue<sup>74</sup> and even in live cell imaging<sup>75</sup>. Some limitations of using quantum dots, such as reduced permeability and steric hindrance difficulties when binding the targets, mainly caused by their relatively large size compared to conventional probes, as well as their tendency to turn on and off ('blinking')<sup>76</sup> still limit the use of quantum-dot labeled probes for transcript imaging but their attractive photophysical properties suggest a huge potential for single-molecule detection.

Finally, a challenge for the future is to combine transcript imaging approaches with quantitative measurements of other cellular constituents, namely proteins and DNA. While protocols tailored to simultaneously perform RNA FISH and immunofluorescence have been shown to be successful in some cases<sup>56,69</sup>, their generic use is still limited by the variabilities inherent in immunofluorescence. Combination of RNA FISH with immunofluorescence or with GFP protein measurements will facilitate decoupling the relative contributions of transcriptional and translational regulation in cells and tissues, whereas combination with DNA FISH can address the expression variability of different clones within a tumor tissue. Such analysis can provide important insights into the combined regulation of protein expression in complex tissues.

### **Figure captions**

Figure 1 – Common steps in ISH. Dashed boxes are method-specific steps. Hapten coupling requires a secondary light-emitting reaction. Single-molecule FISH requires image enhancement in the form of deconvolution or high-pass filtering of the image stacks, followed by automatic spot localization.

Figure 2 – Probe designs for ISH. (a) Schematic of long probes with randomly distributed labeled nucleotides (top) and *Drosophila* embryo hybridized with probes complementary to the hunchback gene (bottom; reprinted from ref. 77). (b) Schematic of shorter probes, each coupled to 3–5 labeled

nucleotides (top) and fluorescence image, showing single mRNA transcripts in mammalian cells (bottom; reprinted from ref. 26). Scale bar, 5  $\mu\text{m}$ . (c) Schematic of multiple singly labeled shorter probes (top) and *C. elegans* embryos hybridized with probes complementary to the *end-1* gene (bottom; reprinted from ref. 31). Scale bar, 5  $\mu\text{m}$ . (d) Illustration of rolling-circle amplification (RCA) of padlock probes<sup>34</sup>; fluorescent probes bind their complementary sequences in the rolling circle product (RCP) (left). Fluorescence image (right) illustrates detection of *ERBB2* (also known as *HER2*; red), *MYC* (yellow) and *ACTB* (blue) transcripts in primary human fibroblast cultures (reprinted from ref. 34). Scale bar, 20  $\mu\text{m}$ . (e) Schematic of a branched DNA detection method using four probe sets. Multiple label probes bind multiple amplifier molecules that are in turn bound to a pre-amplifier molecule that binds the transcript-specific probe pairs. Image of *ERBB2* (green) and 18S rRNA (red) in SKBR3 cells (right; reprinted from <http://www.panomix.com/>).

**Table 1 – Technologies for single-molecule transcript imaging**

Method	Single - molecule resolution	Live-cell imaging	Advantages	Disadvantages	References
Traditional FISH	no	no	Probes can be designed for a wide range of genes	Limited sensitivity and specificity; impossible to detect very short transcripts	15
Probes labeled with multiple fluorophores	yes	no	Generic approach suited for most transcripts; established protocols exist	High variability in the number of probes bound to target	21-26,39
Multiple probes labeled with single fluorophores	yes	no	Generic approach suited for most transcripts, low variability in the number of probes bound to target; commercially available labeled probes	Limited in the detection of very short transcripts	28,30,31 <a href="http://info.biosearchtech.com/singlemoleculefish">http://info.biosearchtech.com/singlemoleculefish</a>
Rolling circle amplification of Padlock	yes	no	Can detect very short probes and discern between single nucleotide	Protocol requires reverse transcription	34

probes			differences	and amplification steps	
Branched DNA probes	yes	no	Commercially available labeled probes	Protocol requires amplification	<sup>38</sup> www.panomics.com
Quantum dot-labeled probes	yes	yes	Bright and photostable; huge spectral range	Large physical size detrimental to target binding and can cause cell penetration problems	70-75
Sub-diffraction microscopy	yes	yes	Achieves a spatial resolution of 20 nanometers	Slow, expensive instrumentation	64-68
MS2-GFP	yes	yes	No need for external interventions (for example microinjection); yields spatial information	Requires generating transgenes, mRNA tend to form clumps	43,44,78
Molecular beacons	yes	yes	High specificity with no clumping; labeled probes are commercially available	Require micro-injecting probes	52-54

### **Acknowledgments**

We thank Stefan Semrau, Jan Philipp Junker, Shankar Mukherji and Anat Lavi-Itzkovitz for valuable comments. This work was supported by the NIH/NCI Physical Sciences Oncology Center at MIT (U54CA143874) and a NIH Pioneer award (1DP1OD003936) to A.v.O. S.I. acknowledges support from EMBO, the Human Frontiers Science Program and the Machiah Foundation.

### **References**

1. M. B. Eisen, P. T. Spellman, P. O. Brown et al., *Proc Natl Acad Sci U S A* 95 (25), 14863 (1998).
2. R. A. Heller, M. Schena, A. Chai et al., *Proc Natl Acad Sci U S A* 94 (6), 2150 (1997).
3. T. Sorlie, C. M. Perou, R. Tibshirani et al., *Proc Natl Acad Sci U S A* 98 (19), 10869 (2001).
4. L. J. van 't Veer, H. Dai, M. J. van de Vijver et al., *Nature* 415 (6871), 530 (2002).
5. J. M. Levsky and R. H. Singer, *Trends Cell Biol* 13 (1), 4 (2003).
6. A. Raj and A. van Oudenaarden, *Cell* 135 (2), 216 (2008).

7. C. Lengauer, K. W. Kinzler, and B. Vogelstein, *Nature* 386 (6625), 623 (1997).
8. S. Y. Park, M. Gonen, H. J. Kim et al., *J Clin Invest* 120 (2), 636 (2010).
9. N. Navin, A. Krasnitz, L. Rodgers et al., *Genome Res* 20 (1), 68 (2010).
10. T. Reya, S. J. Morrison, M. F. Clarke et al., *Nature* 414 (6859), 105 (2001).
11. M. R. Emmert-Buck, R. F. Bonner, P. D. Smith et al., *Science* 274 (5289), 998 (1996).
12. W. M. Freeman, S. J. Walker, and K. E. Vrana, *Biotechniques* 26 (1), 112 (1999).
13. L. Warren, D. Bryder, I. L. Weissman et al., *Proc Natl Acad Sci U S A* 103 (47), 17807 (2006).
14. L. G. van der Flier, M. E. van Gijn, P. Hatzis et al., *Cell* 136 (5), 903 (2009).
15. J. M. Levisky and R. H. Singer, *J Cell Sci* 116 (Pt 14), 2833 (2003).
16. M. van der Ploeg, *Eur J Histochem* 44 (1), 7 (2000).
17. A. Gregorieff and H. Clevers, *Curr Protoc Stem Cell Biol Chapter 2, Unit 2F 1*.
18. A. Pare, D. Lemons, D. Kosman et al., *Curr Biol* 19 (23), 2037 (2009).
19. J. B. Randolph and A. S. Waggoner, *Nucleic Acids Res* 25 (14), 2923 (1997).
20. D. R. Larson, R. H. Singer, and D. Zenklusen, *Trends Cell Biol* 19 (11), 630 (2009).
21. A. M. Femino, F. S. Fay, K. Fogarty et al., *Science* 280 (5363), 585 (1998).

***This pioneering work demonstrated single-molecule transcript imaging in fixed cells using multiply labeled fluorescent probes.***

22. A. M. Femino, K. Fogarty, L. M. Lifshitz et al., *Methods Enzymol* 361, 245 (2003).
23. H. Maamar, A. Raj, and D. Dubnau, *Science* 317 (5837), 526 (2007).
24. D. Zenklusen, D. R. Larson, and R. H. Singer, *Nat Struct Mol Biol* 15 (12), 1263 (2008).
25. R. Z. Tan and A. van Oudenaarden, *Mol Syst Biol* 6, 358 (2010).
26. A. Raj, C. S. Peskin, D. Tranchina et al., *PLoS Biol* 4 (10), e309 (2006).
27. P. Capodiceci, M. Donovan, H. Buchinsky et al., *Nat Methods* 2 (9), 663 (2005).
28. A. Raj, P. van den Bogaard, S. A. Rifkin et al., *Nat Methods* 5 (10), 877 (2008).

***Single-molecule transcript imaging in cells and in fixed tissues was achieved using multiple singly labeled probes.***

29. T. L. To and N. Maheshri, *Science* 327 (5969), 1142 (2010).
30. A. M. Khalil, M. Guttman, M. Huarte et al., *Proc Natl Acad Sci U S A* 106 (28), 11667 (2009).
31. A. Raj, S. A. Rifkin, E. Andersen et al., *Nature* 463 (7283), 913 (2010).
32. Y. Taniguchi, P. J. Choi, G. W. Li et al., *Science* 329 (5991), 533 (2010).
33. J. Lu and A. Tsourkas, *Nucleic Acids Res* 37 (14), e100 (2009).
34. C. Larsson, I. Grundberg, O. Soderberg et al., *Nat Methods* 7 (5), 395 (2010).

***This work introduced a sensitive single-molecule transcript imaging technique that utilizes padlock probes and in situ target-primed rolling-circle amplification in cells and tissues to specifically detect targets with a single nucleotide difference.***

35. P. Wittung, P. E. Nielsen, O. Buchardt et al., *Nature* 368 (6471), 561 (1994).
36. P. M. Lansdorp, N. P. Verwoerd, F. M. van de Rijke et al., *Hum Mol Genet* 5 (5), 685 (1996).
37. W. P. Kloosterman, E. Wienholds, E. de Bruijn et al., *Nat Methods* 3 (1), 27 (2006).
38. A. N. Player, L. P. Shen, D. Kenny et al., *J Histochem Cytochem* 49 (5), 603 (2001).
39. J. M. Levisky, S. M. Shenoy, R. C. Pezo et al., *Science* 297 (5582), 836 (2002).

***This work substantially increased the number of targets that can be simultaneously imaged by using combination of probes labeled with spectrally distinct fluorophores.***

40. P. M. Nederlof, S. van der Flier, J. Vrolijk et al., *Cytometry* 13 (8), 839 (1992).
41. P. M. Nederlof, S. van der Flier, J. Wiegant et al., *Cytometry* 11 (1), 126 (1990).
42. H. Y. Park, A. R. Buxbaum, and R. H. Singer, *Methods Enzymol* 472, 387.
43. I. Golding, J. Paulsson, S. M. Zawilski et al., *Cell* 123 (6), 1025 (2005).
44. E. Bertrand, P. Chartrand, M. Schaefer et al., *Mol Cell* 2 (4), 437 (1998).

**This work introduced the MS2 method for single mRNA detection in live cells.**

45. D. Fusco, N. Accornero, B. Lavoie et al., *Curr Biol* 13 (2), 161 (2003).
46. Y. Shav-Tal, X. Darzacq, S. M. Shenoy et al., *Science* 304 (5678), 1797 (2004).
47. L. Haim, G. Zipor, S. Aronov et al., *Nat Methods* 4 (5), 409 (2007).
48. O. Rackham and C. M. Brown, *Embo J* 23 (16), 3346 (2004).
49. S. Tyagi, *Nat Methods* 6 (5), 331 (2009).
50. M. Valencia-Burton, R. M. McCullough, C. R. Cantor et al., *Nat Methods* 4 (5), 421 (2007).
51. T. Ozawa, Y. Natori, M. Sato et al., *Nat Methods* 4 (5), 413 (2007).
52. S. Tyagi and F. R. Kramer, *Nat Biotechnol* 14 (3), 303 (1996).

**This work introduced molecular beacons for the detection of single mRNA in live cells.**

53. D. Y. Vargas, A. Raj, S. A. Marras et al., *Proc Natl Acad Sci U S A* 102 (47), 17008 (2005).
54. D. P. Bratu, B. J. Cha, M. M. Mhlanga et al., *Proc Natl Acad Sci U S A* 100 (23), 13308 (2003).
55. A. W. Lifland, C. Zurla, and P. J. Santangelo, *Bioconjug Chem* (2010).
56. P. J. Santangelo, A. W. Lifland, P. Curt et al., *Nat Methods* 6 (5), 347 (2009).
57. R. E. Thompson, D. R. Larson, and W. W. Webb, *Biophys J* 82 (5), 2775 (2002).
58. D. Grunwald, R. H. Singer, and K. Czaplinski, *Methods Enzymol* 448, 553 (2008).
59. J. G. McNally, T. Karpova, J. Cooper et al., *Methods* 19 (3), 373 (1999).
60. C.S. Smith, N. Joseph, B. Rieger et al., *Nat Methods* 7 (5), 373 (2010).
61. L. G. van der Flier and H. Clevers, *Annu Rev Physiol* 71, 241 (2009).
62. U. Alon, *Nat Rev Genet* 8 (6), 450 (2007).
63. G. Bassell and R. H. Singer, *Curr Opin Cell Biol* 9 (1), 109 (1997).
64. T. A. Klar, S. Jakobs, M. Dyba et al., *Proc Natl Acad Sci U S A* 97 (15), 8206 (2000).
65. M. J. Rust, M. Bates, and X. Zhuang, *Nat Methods* 3 (10), 793 (2006).
66. E. Betzig, G. H. Patterson, R. Sougrat et al., *Science* 313 (5793), 1642 (2006). 67. S. W. Hell, *Science* 316 (5828), 1153 (2007).
68. H. Shroff, C. G. Galbraith, J. A. Galbraith et al., *Nat Methods* 5 (5), 417 (2008).
69. G. J. Bassell, H. Zhang, A. L. Byrd et al., *J Neurosci* 18 (1), 251 (1998).
70. Y. Choi, H. P. Kim, S. M. Hong et al., *Small* 5 (18), 2085 (2009).
71. P. Chan, T. Yuen, F. Ruf et al., *Nucleic Acids Res* 33 (18), e161 (2005).
72. J. Gonzalez-Maeso, R. L. Ang, T. Yuen et al., *Nature* 452 (7183), 93 (2008).
73. A. M. Smith, H. Duan, A. M. Mohs et al., *Adv Drug Deliv Rev* 60 (11), 1226 (2008).
74. E. Tholouli, J. A. Hoyland, D. Di Vizio et al., *Biochem Biophys Res Commun* 348 (2), 628 (2006).
75. Y. Ishihama and T. Funatsu, *Biochem Biophys Res Commun* 381 (1), 33 (2009).
76. T. T. Weil, R. M. Parton, and I. Davis, *Trends Cell Biol* 20 (7), 380.
77. D. Tautz and C. Pfeifle, *Chromosoma* 98 (2), 81 (1989).
78. K. S. Bloom, D. L. Beach, P. Maddox et al., *Methods Cell Biol* 61, 369 (1999).